

Technical Report for

Providence Engineering

Valero-CAMS, Baton Rouge, LA

712-001

Accutest Job Number: JB58159

Sampling Date: 01/10/14

Report to:

Providence Engineering

kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: **11**



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

Nancy F. Cole

Nancy Cole
Laboratory Director

Client Service contact: Victoria Pushkova 732-329-0200

Certifications: NJ(12129), NY(10983), CA, CT, DE, FL, IL, IN, KS, KY, LA, MA, MD, MI, MT, NC, OH VAP (CL0056), PA, RI, SC, TN, VA, WV, DoD ELAP (L-A-B L2248)

This report shall not be reproduced, except in its entirety, without the written approval of Accutest Laboratories.
Test results relate only to samples analyzed.

Table of Contents

-1-

Section 1: Sample Summary	3
Section 2: Sample Results	4
2.1: JB58159-1: CAMS 146	5
Section 3: Misc. Forms	8
3.1: Chain of Custody	9
3.2: Summa Canister and Flow Controller Log	11



Sample Summary

Providence Engineering

Job No: JB58159

Valero-CAMS, Baton Rouge, LA
Project No: 712-001

Sample Number	Collected		Received	Matrix		Client Sample ID
	Date	Time By		Code	Type	
JB58159-1	01/10/14	13:00 KH	01/20/14	AIR	Ambient Air Grab	CAMS 146

Sample Results

Report of Analysis

Report of Analysis

Page 1 of 3

Client Sample ID:	CAMS 146	Date Sampled:	01/10/14
Lab Sample ID:	JB58159-1	Date Received:	01/20/14
Matrix:	AIR - Ambient Air Grab	Summa ID:	A1026
Method:	TO-15	Percent Solids:	n/a
Project:	Valero-CAMS, Baton Rouge, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3W38495.D	1	01/24/14	YMH	n/a	n/a	V3W1469
Run #2	3W38516.D	1	01/25/14	YMH	n/a	n/a	V3W1470

Run #	Initial Volume
Run #1	400 ml
Run #2	200 ml

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
67-64-1	58.08	Acetone	65.1 ^a	0.40	0.068	ppbv		155 ^a	0.95	0.16	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.020	ppbv		ND	0.44	0.044	ug/m3
71-43-2	78.11	Benzene	0.33	0.20	0.021	ppbv		1.1	0.64	0.067	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.025	ppbv		ND	1.3	0.17	ug/m3
75-25-2	252.8	Bromoform	ND	0.20	0.022	ppbv		ND	2.1	0.23	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.017	ppbv		ND	0.78	0.066	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.014	ppbv		ND	0.87	0.061	ug/m3
100-44-7	126	Benzyl Chloride	ND	0.20	0.025	ppbv		ND	1.0	0.13	ug/m3
75-15-0	76.14	Carbon disulfide	ND	0.20	0.017	ppbv		ND	0.62	0.053	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.025	ppbv		ND	0.92	0.12	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.020	ppbv		ND	0.53	0.053	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.019	ppbv		ND	0.98	0.093	ug/m3
74-87-3	50.49	Chloromethane	0.49	0.20	0.034	ppbv		1.0	0.41	0.070	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.028	ppbv		ND	0.63	0.088	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.020	ppbv		ND	1.0	0.10	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	0.011	ppbv		ND	1.3	0.069	ug/m3
110-82-7	84.16	Cyclohexane	0.56	0.20	0.058	ppbv		1.9	0.69	0.20	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.016	ppbv		ND	0.81	0.065	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.021	ppbv		ND	0.79	0.083	ug/m3
106-93-4	187.9	1,2-Dibromoethane	ND	0.20	0.027	ppbv		ND	1.5	0.21	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.016	ppbv		ND	0.81	0.065	ug/m3
78-87-5	113	1,2-Dichloropropane	0.73	0.20	0.040	ppbv		3.4	0.92	0.18	ug/m3
123-91-1	88.12	1,4-Dioxane	0.52	0.20	0.060	ppbv		1.9	0.72	0.22	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.50	0.20	0.015	ppbv		2.5	0.99	0.074	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.029	ppbv		ND	1.7	0.25	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.015	ppbv		ND	0.79	0.059	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.028	ppbv		ND	0.79	0.11	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.019	ppbv		ND	0.91	0.086	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.025	ppbv		ND	1.2	0.15	ug/m3
95-50-1	147	o-Dichlorobenzene	ND	0.20	0.029	ppbv		ND	1.2	0.17	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.022	ppbv		ND	1.2	0.13	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.021	ppbv		ND	0.91	0.095	ug/m3

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	CAMS 146	Date Sampled:	01/10/14
Lab Sample ID:	JB58159-1	Date Received:	01/20/14
Matrix:	AIR - Ambient Air Grab	Summa ID:	A1026
Method:	TO-15	Percent Solids:	n/a
Project:	Valero-CAMS, Baton Rouge, LA		

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
64-17-5	46.07	Ethanol	4.5	0.50	0.19	ppbv		8.5	0.94	0.36	ug/m3
100-41-4	106.2	Ethylbenzene	0.13	0.20	0.020	ppbv	J	0.56	0.87	0.087	ug/m3
141-78-6	88	Ethyl Acetate	ND	0.20	0.057	ppbv		ND	0.72	0.21	ug/m3
622-96-8	120.2	4-Ethyltoluene	ND	0.20	0.015	ppbv		ND	0.98	0.074	ug/m3
76-13-1	187.4	Freon 113	0.49	0.20	0.021	ppbv		3.8	1.5	0.16	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.021	ppbv		ND	1.4	0.15	ug/m3
142-82-5	100.2	Heptane	0.14	0.20	0.020	ppbv	J	0.57	0.82	0.082	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.063	ppbv		ND	2.1	0.67	ug/m3
110-54-3	86.17	Hexane	0.37	0.20	0.016	ppbv		1.3	0.70	0.056	ug/m3
591-78-6	100	2-Hexanone	0.26	0.20	0.025	ppbv		1.1	0.82	0.10	ug/m3
67-63-0	60.1	Isopropyl Alcohol	1.0	0.20	0.039	ppbv		2.5	0.49	0.096	ug/m3
75-09-2	84.94	Methylene chloride	0.66	0.20	0.047	ppbv		2.3	0.69	0.16	ug/m3
78-93-3	72.11	Methyl ethyl ketone	1.6	0.20	0.058	ppbv		4.7	0.59	0.17	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	0.26	0.20	0.029	ppbv		1.1	0.82	0.12	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.017	ppbv		ND	0.72	0.061	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.040	ppbv		ND	0.82	0.16	ug/m3
115-07-1	42	Propylene	ND	0.50	0.031	ppbv		ND	0.86	0.053	ug/m3
100-42-5	104.1	Styrene	ND	0.20	0.020	ppbv		ND	0.85	0.085	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.016	ppbv		ND	1.1	0.087	ug/m3
79-34-5	167.9	1,1,2,2-Tetrachloroethane	ND	0.20	0.030	ppbv		ND	1.4	0.21	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.031	ppbv		ND	1.1	0.17	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.079	ppbv		ND	1.5	0.59	ug/m3
95-63-6	120.2	1,2,4-Trimethylbenzene	ND	0.20	0.017	ppbv		ND	0.98	0.084	ug/m3
108-67-8	120.2	1,3,5-Trimethylbenzene	ND	0.20	0.015	ppbv		ND	0.98	0.074	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	ND	0.20	0.021	ppbv		ND	0.93	0.098	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	0.66	0.20	0.044	ppbv		2.0	0.61	0.13	ug/m3
127-18-4	165.8	Tetrachloroethylene	0.048	0.040	0.029	ppbv		0.33	0.27	0.20	ug/m3
109-99-9	72.11	Tetrahydrofuran	0.18	0.20	0.045	ppbv	J	0.53	0.59	0.13	ug/m3
108-88-3	92.14	Toluene	0.46	0.20	0.020	ppbv		1.7	0.75	0.075	ug/m3
79-01-6	131.4	Trichloroethylene	ND	0.040	0.019	ppbv		ND	0.21	0.10	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.26	0.20	0.014	ppbv		1.5	1.1	0.079	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.017	ppbv		ND	0.51	0.043	ug/m3
108-05-4	86	Vinyl Acetate	ND	0.20	0.058	ppbv		ND	0.70	0.20	ug/m3
	106.2	m,p-Xylene	0.37	0.20	0.032	ppbv		1.6	0.87	0.14	ug/m3
95-47-6	106.2	o-Xylene	0.15	0.20	0.019	ppbv	J	0.65	0.87	0.083	ug/m3
1330-20-7	106.2	Xylenes (total)	0.52	0.20	0.019	ppbv		2.3	0.87	0.083	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	95%	94%	65-128%

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID: CAMS 146		Date Sampled: 01/10/14
Lab Sample ID: JB58159-1		Date Received: 01/20/14
Matrix: AIR - Ambient Air Grab	Summa ID: A1026	Percent Solids: n/a
Method: TO-15		
Project: Valero-CAMS, Baton Rouge, LA		

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
---------	----	----------	--------	----	-----	-------	---	--------	----	-----	-------

(a) Result is from Run# 2

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody
- Summa Canister and Flow Controller Log

AIR

CHAIN OF CUSTODY

Air Sampling Field Data Sheet



2235 US Highway 130, Dayton, NJ 08810
V: 732.329.0200 F: 732.329.3499 www.accutest.com

FED-EX Tracking #
79763331518
Lab Quote #

Bottle Order Control #
Lab Job # JB58159

PAGE 1 OF 1

Client / Reporting Information		Project Information		Weather Parameters		Requested Analysis	
Company Name: Providence Engr		Project Name: Valero Refinery		Temperature (Fahrenheit)			
Address: 1201 Main St		Street:		Start: Maximum:			
City: BIR State: LA Zip: 70802		City: Meroux State: CA		Stop: Minimum:			
Project Contact: Paul Hollis e: providenceeng.com		Project #: 712-001		Atmospheric Pressure (inches of Hg)			
Phone #: 225 766-7400 Fax #: -7440		Client Purchase Order #		Start: Maximum:		TO-15	
Sampler(s) Name(s): Karen Hudson				Stop: Minimum:			
Other weather comment:							

Lab Sample #	Field ID / Point of Collection	Air Type				Sampling Equipment Info					Start Sampling Information					Stop Sampling Information				
		Indoor(I)	Soil Vap(SV)	Ambient(A)		Canister Serial #	Canister Size SL or TL	Flow Controller Serial #	Date	Time (24hr clock)	Canister Pressure (Hg)	Interior Temp (F)	Sampler Init.	Date	Time (24hr clock)	Canister Pressure (Hg)	Interior Temp (F)	Sampler Init.		
1	CAMS 146	A				A1026	6L	-	1-9	1300	0.11	75	ket	1-10	1300	11.56	75	ket		

Turnaround Time (Business days)		Data Deliverable Information		Comments / Remarks	
Standard - 15 Days	<input checked="" type="checkbox"/>	Approved By: _____	All NJDEP TO-15 is mandatory Full T1	SUMMARY Received at Baton Rouge Service Center	
10 Day	<input type="checkbox"/>	Date: _____	Comm A		
5 Day	<input type="checkbox"/>		Comm B		
3 Day	<input type="checkbox"/>		Reduced T2		
2 Day	<input type="checkbox"/>		Full T1		
1 Day	<input type="checkbox"/>		Other: _____		
Other	<input type="checkbox"/>				

Sample Custody must be documented below each time samples change possession, including courier delivery.

Relinquished by Laboratory: 1	Date Time:	Received By: [Signature]	Relinquished By: 2	Date Time: 1/16/14 12:05	Received By: [Signature]
Relinquished by: FEDEX	Date Time: 1-20-14	Received By: [Signature]	Relinquished By: 3		Received By: [Signature]
Relinquished by: 5	Date Time:	Received By: [Signature]	Custody Seal #		

JB58159: Chain of Custody

Page 1 of 2



Accutest Laboratories Sample Receipt Summary

Accutest Job Number: JB58159 **Client:** _____ **Project:** _____
Date / Time Received: 1/20/2014 **Delivery Method:** _____ **Airbill #s:** _____

Cooler Temps (Initial/Adjusted):

<u>Cooler Security</u>	<u>Y or N</u>		<u>Y or N</u>	
1. Custody Seals Present:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. COC Present:	<input checked="" type="checkbox"/> <input type="checkbox"/>
2. Custody Seals Intact:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. Smpl Dates/Time OK	<input checked="" type="checkbox"/> <input type="checkbox"/>

<u>Cooler Temperature</u>	<u>Y or N</u>	
1. Temp criteria achieved:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2. Cooler temp verification:	_____	
3. Cooler media:	_____	
4. No. Coolers:	0	

<u>Quality Control Preservation</u>	<u>Y</u>	<u>or</u>	<u>N</u>	<u>N/A</u>
1. Trip Blank present / cooler:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
2. Trip Blank listed on COC:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
3. Samples preserved properly:	<input checked="" type="checkbox"/>	<input type="checkbox"/>		
4. VOCs headspace free:	<input type="checkbox"/>	<input type="checkbox"/>		<input checked="" type="checkbox"/>

<u>Sample Integrity - Documentation</u>	<u>Y or N</u>	
1. Sample labels present on bottles:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2. Container labeling complete:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
3. Sample container label / COC agree:	<input checked="" type="checkbox"/>	<input type="checkbox"/>

<u>Sample Integrity - Condition</u>	<u>Y or N</u>	
1. Sample recvd within HT:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2. All containers accounted for:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
3. Condition of sample:	Intact	

<u>Sample Integrity - Instructions</u>	<u>Y</u>	<u>or</u>	<u>N</u>	<u>N/A</u>
1. Analysis requested is clear:	<input checked="" type="checkbox"/>	<input type="checkbox"/>		
2. Bottles received for unspecified tests	<input type="checkbox"/>	<input checked="" type="checkbox"/>		
3. Sufficient volume recvd for analysis:	<input checked="" type="checkbox"/>	<input type="checkbox"/>		
4. Compositing instructions clear:	<input type="checkbox"/>	<input type="checkbox"/>		<input checked="" type="checkbox"/>
5. Filtering instructions clear:	<input type="checkbox"/>	<input type="checkbox"/>		<input checked="" type="checkbox"/>

Comments

Accutest Laboratories
V: 732.329.0200

2235 US Highway 130
F: 732.329.3499

Dayton, New Jersey
www.accutest.com

3.1
3

Summa Canister and Flow Controller Log

Job Number: JB58159
Account: PROVLABR Providence Engineering
Project: Valero-CAMS, Baton Rouge, LA
Received: 01/20/14

32
3

SUMMA CANISTERS													
Shipping							Receiving						
Summa ID	L	Vac " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
A1026	6	29.4	12/18/13	RC	CP6665	5W1918.D	JB58159-1	01/23/14	FZ	6.5			1

Accutest Bottle Order(s):
 VP-12/16/2013-16

Prep Date **Room Temp(F)** **Bar Pres "Hg**
 12/18/13 70 29.92