

**Technical Report for**

**Providence Engineering**

Valero-CAMS, Baton Rouge, LA

712-001

Accutest Job Number: JB56495

Sampling Date: 12/23/13

**Report to:**

Providence Engineering

kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: **10**



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

*Nancy F. Cole*

**Nancy Cole**  
**Laboratory Director**

**Client Service contact: Victoria Pushkova 732-329-0200**

Certifications: NJ(12129), NY(10983), CA, CT, DE, FL, IL, IN, KS, KY, LA, MA, MD, MI, MT, NC, OH VAP (CL0056), PA, RI, SC, TN, VA, WV, DoD ELAP (L-A-B L2248)

This report shall not be reproduced, except in its entirety, without the written approval of Accutest Laboratories.  
Test results relate only to samples analyzed.

# Table of Contents

-1-

<b>Section 1: Sample Summary</b> .....	<b>3</b>
<b>Section 2: Sample Results</b> .....	<b>4</b>
<b>2.1: JB56495-1: CAMS 143</b> .....	<b>5</b>
<b>Section 3: Misc. Forms</b> .....	<b>7</b>
<b>3.1: Chain of Custody</b> .....	<b>8</b>
<b>3.2: Summa Canister and Flow Controller Log</b> .....	<b>10</b>



## Sample Summary

**Providence Engineering**

**Job No: JB56495**

**Valero-CAMS, Baton Rouge, LA**  
**Project No: 712-001**

Sample Number	Collected		Received	Matrix		Client Sample ID
	Date	Time By		Code	Type	
JB56495-1	12/23/13	13:00 KH	12/27/13	AIR	Ambient Air Grab	CAMS 143

**Sample Results**

---

**Report of Analysis**

---

## Report of Analysis

Client Sample ID:	CAMS 143	Date Sampled:	12/23/13
Lab Sample ID:	JB56495-1	Date Received:	12/27/13
Matrix:	AIR - Ambient Air Grab Summa ID: A354	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	5W2153.D	1	01/02/14	ML	n/a	n/a	V5W86
Run #2							

Run #	Initial Volume
Run #1	400 ml
Run #2	

## VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
67-64-1	58.08	Acetone	8.2	0.20	0.034	ppbv		19	0.48	0.081	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.020	ppbv		ND	0.44	0.044	ug/m3
71-43-2	78.11	Benzene	0.15	0.20	0.021	ppbv	J	0.48	0.64	0.067	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.025	ppbv		ND	1.3	0.17	ug/m3
75-25-2	252.8	Bromoform	ND	0.20	0.022	ppbv		ND	2.1	0.23	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.017	ppbv		ND	0.78	0.066	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.014	ppbv		ND	0.87	0.061	ug/m3
100-44-7	126	Benzyl Chloride	ND	0.20	0.025	ppbv		ND	1.0	0.13	ug/m3
75-15-0	76.14	Carbon disulfide	0.14	0.20	0.017	ppbv	J	0.44	0.62	0.053	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.025	ppbv		ND	0.92	0.12	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.020	ppbv		ND	0.53	0.053	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.019	ppbv		ND	0.98	0.093	ug/m3
74-87-3	50.49	Chloromethane	0.69	0.20	0.034	ppbv		1.4	0.41	0.070	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.028	ppbv		ND	0.63	0.088	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.020	ppbv		ND	1.0	0.10	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	0.011	ppbv		ND	1.3	0.069	ug/m3
110-82-7	84.16	Cyclohexane	ND	0.20	0.058	ppbv		ND	0.69	0.20	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.016	ppbv		ND	0.81	0.065	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.021	ppbv		ND	0.79	0.083	ug/m3
106-93-4	187.9	1,2-Dibromoethane	ND	0.20	0.027	ppbv		ND	1.5	0.21	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.016	ppbv		ND	0.81	0.065	ug/m3
78-87-5	113	1,2-Dichloropropane	ND	0.20	0.040	ppbv		ND	0.92	0.18	ug/m3
123-91-1	88.12	1,4-Dioxane	ND	0.20	0.060	ppbv		ND	0.72	0.22	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.48	0.20	0.015	ppbv		2.4	0.99	0.074	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.029	ppbv		ND	1.7	0.25	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.015	ppbv		ND	0.79	0.059	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.028	ppbv		ND	0.79	0.11	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.019	ppbv		ND	0.91	0.086	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.025	ppbv		ND	1.2	0.15	ug/m3
95-50-1	147	o-Dichlorobenzene	ND	0.20	0.029	ppbv		ND	1.2	0.17	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.022	ppbv		ND	1.2	0.13	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.021	ppbv		ND	0.91	0.095	ug/m3

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

## Report of Analysis

Client Sample ID:	CAMS 143	Date Sampled:	12/23/13
Lab Sample ID:	JB56495-1	Date Received:	12/27/13
Matrix:	AIR - Ambient Air Grab Summa ID: A354	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

## VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
64-17-5	46.07	Ethanol	2.4	0.50	0.19	ppbv		4.5	0.94	0.36	ug/m3
100-41-4	106.2	Ethylbenzene	0.086	0.20	0.020	ppbv	J	0.37	0.87	0.087	ug/m3
141-78-6	88	Ethyl Acetate	ND	0.20	0.057	ppbv		ND	0.72	0.21	ug/m3
622-96-8	120.2	4-Ethyltoluene	ND	0.20	0.015	ppbv		ND	0.98	0.074	ug/m3
76-13-1	187.4	Freon 113	2.5	0.20	0.021	ppbv		19	1.5	0.16	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.021	ppbv		ND	1.4	0.15	ug/m3
142-82-5	100.2	Heptane	0.10	0.20	0.020	ppbv	J	0.41	0.82	0.082	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.063	ppbv		ND	2.1	0.67	ug/m3
110-54-3	86.17	Hexane	0.75	0.20	0.016	ppbv		2.6	0.70	0.056	ug/m3
591-78-6	100	2-Hexanone	0.41	0.20	0.025	ppbv		1.7	0.82	0.10	ug/m3
67-63-0	60.1	Isopropyl Alcohol	0.52	0.20	0.039	ppbv		1.3	0.49	0.096	ug/m3
75-09-2	84.94	Methylene chloride	2.5	0.20	0.047	ppbv		8.7	0.69	0.16	ug/m3
78-93-3	72.11	Methyl ethyl ketone	1.8	0.20	0.058	ppbv		5.3	0.59	0.17	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	0.28	0.20	0.029	ppbv		1.1	0.82	0.12	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.017	ppbv		ND	0.72	0.061	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.040	ppbv		ND	0.82	0.16	ug/m3
115-07-1	42	Propylene	1.1	0.50	0.031	ppbv		1.9	0.86	0.053	ug/m3
100-42-5	104.1	Styrene	0.16	0.20	0.020	ppbv	J	0.68	0.85	0.085	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.016	ppbv		ND	1.1	0.087	ug/m3
79-34-5	167.9	1,1,2,2-Tetrachloroethane	ND	0.20	0.030	ppbv		ND	1.4	0.21	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.031	ppbv		ND	1.1	0.17	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.079	ppbv		ND	1.5	0.59	ug/m3
95-63-6	120.2	1,2,4-Trimethylbenzene	0.11	0.20	0.017	ppbv	J	0.54	0.98	0.084	ug/m3
108-67-8	120.2	1,3,5-Trimethylbenzene	ND	0.20	0.015	ppbv		ND	0.98	0.074	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	ND	0.20	0.021	ppbv		ND	0.93	0.098	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	ND	0.20	0.044	ppbv		ND	0.61	0.13	ug/m3
127-18-4	165.8	Tetrachloroethylene	ND	0.040	0.029	ppbv		ND	0.27	0.20	ug/m3
109-99-9	72.11	Tetrahydrofuran	ND	0.20	0.045	ppbv		ND	0.59	0.13	ug/m3
108-88-3	92.14	Toluene	0.57	0.20	0.020	ppbv		2.1	0.75	0.075	ug/m3
79-01-6	131.4	Trichloroethylene	ND	0.040	0.019	ppbv		ND	0.21	0.10	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.23	0.20	0.014	ppbv		1.3	1.1	0.079	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.017	ppbv		ND	0.51	0.043	ug/m3
108-05-4	86	Vinyl Acetate	ND	0.20	0.058	ppbv		ND	0.70	0.20	ug/m3
	106.2	m,p-Xylene	0.28	0.20	0.032	ppbv		1.2	0.87	0.14	ug/m3
95-47-6	106.2	o-Xylene	0.098	0.20	0.019	ppbv	J	0.43	0.87	0.083	ug/m3
1330-20-7	106.2	Xylenes (total)	0.38	0.20	0.019	ppbv		1.7	0.87	0.083	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	100%		65-128%

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

## Misc. Forms

---

### Custody Documents and Other Forms

---

**Includes the following where applicable:**

- Chain of Custody
- Summa Canister and Flow Controller Log

# CHAIN OF CUSTODY

## Air Sampling Field Data Sheet



2235 US Highway 130, Dayton, NJ 08810  
 V: 732.329.0200 F: 732.329.3499 www.accutest.com

FEDEX Tracking # 1475 8751 1571  
 Bottle Order Control #  
 Lab Quote # JB56495

PAGE 1 OF 1

Client / Reporting Information				Project Information				Weather Parameters				Requested Analysis				
Company Name: Providence Engr				Project Name: Valero Refining				Temperature (Fahrenheit)				Requested Analysis				
Address: 1201 Main St				Street:				Start: Maximum:								
City: Bz State: LA Zip: 70802				City: Meraux LA State:				Stop: Minimum:								
Project Contact: Paul Hollis E-prov.denceeng.com				Project #: 712-001				Atmospheric Pressure (inches of Hg)								
Phone #: 225-766-7400 Fax #: -7440				Client Purchase Order #				Start: Maximum:				Requested Analysis				
Sampler(s) Name(s): K Hudson				Other weather comment:				Stop: Minimum:								
Lab Sample #	Field ID / Point of Collection	Air Type	Sampling Equipment Info			Start Sampling Information					Stop Sampling Information					
		Indoor (I) Soil Vap (SV) Ambient (A)	Canister Serial #	Canister Size EL or 1L	Flow Controller Serial #	Date	Time (24hr clock)	Canister Pressure (Hig)	Interior Temp (F)	Sampler Init.	Date	Time (24hr clock)	Canister Pressure (Hig)	Interior Temp (F)	Sampler Init.	
-1	CAMS 143	A	A354	6L	-	12-22	1300	0.03	75	KH	12-23	1300	11.32	75	KH	X
Turnaround Time (Business days)				Data Deliverable Information				Comments / Remarks				Requested Analysis				
Standard - 15 Days 10 Day 5 Day 3 Day 2 Day 1 Day Other				Approved By: _____ Date: _____				All NJDEP TO-15 is mandatory Full T1 Comm A Comm B Reduced T2 Full T1 Other: _____						Received at Baton Rouge Service Center Summa		
Sample Custody must be documented below each time samples change possession, including courier delivery.																
Relinquished by Laboratory: 1	Date Time:	Received By:	Relinquished By:	Date Time:	Received By:	Relinquished by:	Date Time:	Received By:	Relinquished By:	Date Time:	Received By:	Relinquished by:	Date Time:	Received By:	Relinquished By:	Date Time:
3	12/27/13 15:30	3 FedEx	4 FedEx	12/26/13 11:12	2	3	12/27/13 9:15	4	4	4	4	5	5	5	5	5

31  
3

10-15

JK





# Summa Canister and Flow Controller Log

**Job Number:** JB56495  
**Account:** PROVLABR Providence Engineering  
**Project:** Valero-CAMS, Baton Rouge, LA  
**Received:** 12/27/13

32  
3

SUMMA CANISTERS													
Shipping							Receiving						
Summa ID	Vac L	Date " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
A354	6	29.4	11/22/13	RC	CP6574	5W964.D	JB56495-1	12/27/13	ML	7			1

**Accutest Bottle Order(s):**  
 VP-11/22/2013-4

**Prep Date**      **Room Temp(F)**      **Bar Pres "Hg**  
 11/22/13          70                                  29.92