

Effective January 1, 2016, SGS has acquired all of the assets of Accutest Laboratories and will continue to operate as SGS-Accutest. SGS-Accutest is part of SGS, the world's leading inspection, verification, testing and certification company.

## Technical Report for

### Providence Engineering

Valero-CAMS, Baton Rouge, LA

712-001

Accutest Job Number: JC12209

Sampling Date: 12/31/15

#### Report to:

Providence Engineering

kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: **10**



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

A handwritten signature in black ink that reads "Nancy Cole".

**Nancy Cole**  
Laboratory Director

**Client Service contact: Victoria Pushkova 732-329-0200**

Certifications: NJ(12129), NY(10983), CA, CT, DE, FL, IL, IN, KS, KY, LA, MA, MD, MI, MT, NC, OH VAP (CL0056), AK (UST-103), AZ (AZ0786), PA, RI, SC, TN, TX, VA, WV, DoD ELAP (L-A-B L2248)

This report shall not be reproduced, except in its entirety, without the written approval of Accutest Laboratories. Test results relate only to samples analyzed.

# Table of Contents

-1-

<b>Section 1: Sample Summary</b> .....	<b>3</b>
<b>Section 2: Sample Results</b> .....	<b>4</b>
<b>2.1: JC12209-1: CAMS 266</b> .....	<b>5</b>
<b>Section 3: Misc. Forms</b> .....	<b>7</b>
<b>3.1: Chain of Custody</b> .....	<b>8</b>
<b>3.2: Summa Canister and Flow Controller Log</b> .....	<b>10</b>



## Sample Summary

**Providence Engineering**

**Job No: JC12209**

**Valero-CAMS, Baton Rouge, LA**  
**Project No: 712-001**

Sample Number	Collected		Received	Matrix		Client Sample ID
	Date	Time By		Code	Type	
JC12209-1	12/31/15	13:00 KH	01/08/16	AIR	Ambient Air Grab	CAMS 266

**Sample Results**

---

**Report of Analysis**

---

## Report of Analysis

Client Sample ID:	CAMS 266	Date Sampled:	12/31/15
Lab Sample ID:	JC12209-1	Date Received:	01/08/16
Matrix:	AIR - Ambient Air Grab Summa ID: A1194	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	5W15619.D	1.52	01/09/16	TCH	n/a	n/a	V5W626
Run #2							

Run #	Initial Volume
Run #1	608 ml
Run #2	

## VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
67-64-1	58.08	Acetone	3.0	0.20	0.032	ppbv		7.1	0.48	0.076	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.031	ppbv		ND	0.44	0.069	ug/m3
71-43-2	78.11	Benzene	0.17	0.20	0.030	ppbv	J	0.54	0.64	0.096	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.032	ppbv		ND	1.3	0.21	ug/m3
75-25-2	252.8	Bromoform	ND	0.20	0.020	ppbv		ND	2.1	0.21	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.022	ppbv		ND	0.78	0.085	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.020	ppbv		ND	0.87	0.087	ug/m3
100-44-7	126	Benzyl Chloride	ND	0.20	0.026	ppbv		ND	1.0	0.13	ug/m3
75-15-0	76.14	Carbon disulfide	ND	0.20	0.029	ppbv		ND	0.62	0.090	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.032	ppbv		ND	0.92	0.15	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.022	ppbv		ND	0.53	0.058	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.031	ppbv		ND	0.98	0.15	ug/m3
74-87-3	50.49	Chloromethane	0.85	0.20	0.029	ppbv		1.8	0.41	0.060	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.028	ppbv		ND	0.63	0.088	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.033	ppbv		ND	1.0	0.17	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	0.025	ppbv		ND	1.3	0.16	ug/m3
110-82-7	84.16	Cyclohexane	ND	0.20	0.032	ppbv		ND	0.69	0.11	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.031	ppbv		ND	0.81	0.13	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.028	ppbv		ND	0.79	0.11	ug/m3
106-93-4	187.9	1,2-Dibromoethane	ND	0.20	0.035	ppbv		ND	1.5	0.27	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.026	ppbv		ND	0.81	0.11	ug/m3
78-87-5	113	1,2-Dichloropropane	ND	0.20	0.050	ppbv		ND	0.92	0.23	ug/m3
123-91-1	88.12	1,4-Dioxane	ND	0.20	0.063	ppbv		ND	0.72	0.23	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.60	0.20	0.037	ppbv		3.0	0.99	0.18	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.041	ppbv		ND	1.7	0.35	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.020	ppbv		ND	0.79	0.079	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.025	ppbv		ND	0.79	0.099	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.035	ppbv		ND	0.91	0.16	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.028	ppbv		ND	1.2	0.17	ug/m3
95-50-1	147	o-Dichlorobenzene	ND	0.20	0.030	ppbv		ND	1.2	0.18	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.019	ppbv		ND	1.2	0.11	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.020	ppbv		ND	0.91	0.091	ug/m3

ND = Not detected MDL = Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

## Report of Analysis

Client Sample ID:	CAMS 266	Date Sampled:	12/31/15
Lab Sample ID:	JC12209-1	Date Received:	01/08/16
Matrix:	AIR - Ambient Air Grab	Summa ID:	A1194
Method:	TO-15	Percent Solids:	n/a
Project:	Valero-CAMS, Baton Rouge, LA		

## VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
64-17-5	46.07	Ethanol	1.4	0.50	0.17	ppbv		2.6	0.94	0.32	ug/m3
100-41-4	106.2	Ethylbenzene	ND	0.20	0.048	ppbv		ND	0.87	0.21	ug/m3
141-78-6	88	Ethyl Acetate	ND	0.20	0.064	ppbv		ND	0.72	0.23	ug/m3
622-96-8	120.2	4-Ethyltoluene	ND	0.20	0.022	ppbv		ND	0.98	0.11	ug/m3
76-13-1	187.4	Freon 113	ND	0.20	0.027	ppbv		ND	1.5	0.21	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.025	ppbv		ND	1.4	0.17	ug/m3
142-82-5	100.2	Heptane	ND	0.20	0.029	ppbv		ND	0.82	0.12	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.033	ppbv		ND	2.1	0.35	ug/m3
110-54-3	86.17	Hexane	0.20	0.20	0.028	ppbv		0.70	0.70	0.099	ug/m3
591-78-6	100	2-Hexanone	ND	0.20	0.044	ppbv		ND	0.82	0.18	ug/m3
67-63-0	60.1	Isopropyl Alcohol	0.53	0.20	0.12	ppbv		1.3	0.49	0.29	ug/m3
75-09-2	84.94	Methylene chloride	ND	0.20	0.13	ppbv		ND	0.69	0.45	ug/m3
78-93-3	72.11	Methyl ethyl ketone	0.24	0.20	0.049	ppbv		0.71	0.59	0.14	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	ND	0.20	0.027	ppbv		ND	0.82	0.11	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.026	ppbv		ND	0.72	0.094	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.030	ppbv		ND	0.82	0.12	ug/m3
115-07-1	42	Propylene	ND	0.50	0.081	ppbv		ND	0.86	0.14	ug/m3
100-42-5	104.1	Styrene	ND	0.20	0.026	ppbv		ND	0.85	0.11	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.032	ppbv		ND	1.1	0.17	ug/m3
79-34-5	167.9	1,1,2,2-Tetrachloroethane	ND	0.20	0.030	ppbv		ND	1.4	0.21	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.036	ppbv		ND	1.1	0.20	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.044	ppbv		ND	1.5	0.33	ug/m3
95-63-6	120.2	1,2,4-Trimethylbenzene	ND	0.20	0.023	ppbv		ND	0.98	0.11	ug/m3
108-67-8	120.2	1,3,5-Trimethylbenzene	ND	0.20	0.030	ppbv		ND	0.98	0.15	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	ND	0.20	0.021	ppbv		ND	0.93	0.098	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	0.19	0.20	0.050	ppbv	J	0.58	0.61	0.15	ug/m3
127-18-4	165.8	Tetrachloroethylene	ND	0.040	0.024	ppbv		ND	0.27	0.16	ug/m3
109-99-9	72.11	Tetrahydrofuran	ND	0.20	0.043	ppbv		ND	0.59	0.13	ug/m3
108-88-3	92.14	Toluene	0.11	0.20	0.020	ppbv	J	0.41	0.75	0.075	ug/m3
79-01-6	131.4	Trichloroethylene	ND	0.040	0.025	ppbv		ND	0.21	0.13	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.30	0.20	0.020	ppbv		1.7	1.1	0.11	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.032	ppbv		ND	0.51	0.082	ug/m3
108-05-4	86	Vinyl Acetate	ND	0.20	0.055	ppbv		ND	0.70	0.19	ug/m3
	106.2	m,p-Xylene	ND	0.20	0.043	ppbv		ND	0.87	0.19	ug/m3
95-47-6	106.2	o-Xylene	ND	0.20	0.026	ppbv		ND	0.87	0.11	ug/m3
1330-20-7	106.2	Xylenes (total)	ND	0.20	0.026	ppbv		ND	0.87	0.11	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	102%		65-128%

ND = Not detected MDL = Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

## Misc. Forms

---

### Custody Documents and Other Forms

---

**Includes the following where applicable:**

- Chain of Custody
- Summa Canister and Flow Controller Log

AIR



# CHAIN OF CUSTODY

## Air Sampling Field Data Sheet

FED-EX Tracking # 6574 9163 8805 Bottle/Order Control # VP-1212312015-5  
 Lab Quote # Lab Job # JC12209

Client / Reporting Information						Weather Parameters					Requested Analysis				
Company Name <u>Providence Engr</u>			Project Name <u>Valero Refining</u>			Temperature (Fahrenheit)					Standard TO-15 Reporting List				
Address <u>1201 Main St</u>			Street			Start: Maximum:									
City <u>BIR</u> State <u>CA</u> Zip <u>70802</u>			City <u>Meraux</u> State <u>LA</u>			Stop: Minimum:									
Project Contact <u>paullhollos@providenceeng.com</u> E-mail			Project #			Atmospheric Pressure (inches of Hg)									
Phone # <u>(25) 766-7400</u> Fax # <u>-7440</u>			Client Purchase Order #			Start: Maximum:									
Sampler(s) Name(s) <u>K Hudson</u>						Stop: Minimum:									
Other weather comment:															
Lab Sample #	Field ID / Point of Collection	Air Type			Start Sampling Information					Stop Sampling Information					
		Indoor(I) Soil Vap(SV) Ambient(A)	Canister Serial #	Canister Size 6L or 1L	Flow Controller Serial #	Date	Time (24hr clock)	Canister Pressure (*Hg)	Interior Temp (F)	Sampler Init.	Date	Time (24hr clock)	Canister Pressure (*Hg)	Interior Temp (F)	Sampler Init.
<u>-1</u>	<u>CAMS 266</u>	<u>A</u>	<u>A11946L</u>	<u>-</u>	<u>12-30</u>	<u>1300</u>	<u>0.07</u>	<u>75</u>	<u>KH</u>	<u>12-31</u>	<u>1300</u>	<u>0.68</u>	<u>75</u>	<u>KH</u>	<input checked="" type="checkbox"/>
Turnaround Time (Business days)		Data Deliverable Information			Comments / Remarks										
Standard - 15 Days <input checked="" type="checkbox"/>		INITIAL ASSESSMENT <u>HABV</u>			All NJDEP TO-15 is mandatory Full T1										
10 Day		LABEL VERIFICATION <u>BV</u>			Comm A										
5 Day		Date: _____			Comm B										
3 Day					Reduced T2										
2 Day					Full T1										
1 Day					Other:										
Other															
Sample Custody must be documented below each time samples change possession, including courier delivery.															
Relinquished by Label # <u>1</u> <u>Way Mangano</u>		Date Time: <u>12/28/15 14:20</u>		Received By: <u>Fed Ex</u>			Relinquished By: <u>Fed Ex</u>		Date Time: _____		Received By: <u>Summa</u>				
Relinquished by: <u>2</u>		Date Time: _____		Received By: <u>Fed Ex</u>			Relinquished By: <u>Fed Ex</u>		Date Time: <u>1/8/16 9:15</u>		Received By: <u>4</u>				
Relinquished by: <u>3</u>		Date Time: _____		Received By: <u>Fed Ex</u>			Relinquished By: <u>Fed Ex</u>		Date Time: _____		Received By: _____				
Relinquished by: <u>4</u>		Date Time: _____		Received By: _____			Custody Seal # <u>510 Intact</u>		Date Time: _____		Received By: _____				
Relinquished by: <u>5</u>		Date Time: _____		Received By: _____			Custody Seal # _____		Date Time: _____		Received By: _____				

31  
3

JC12209: Chain of Custody

Page 1 of 2



**Accutest Job Number:** JC12209      **Client:** \_\_\_\_\_      **Project:** \_\_\_\_\_  
**Date / Time Received:** 1/8/2016 9:15:00 AM      **Delivery Method:** \_\_\_\_\_      **Airbill #s:** \_\_\_\_\_

Cooler Temps (Raw Measured) °C: \_\_\_\_\_  
 Cooler Temps (Corrected) °C: \_\_\_\_\_

**Cooler Security**

	Y	or	N		Y	or	N
1. Custody Seals Present:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	3. COC Present:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
2. Custody Seals Intact:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	4. Smpl Dates/Time OK	<input checked="" type="checkbox"/>		<input type="checkbox"/>

**Cooler Temperature**

	Y	or	N
1. Temp criteria achieved:	<input type="checkbox"/>		<input type="checkbox"/>
2. Cooler temp verification:	N/A		
3. Cooler media:	N/A		
4. No. Coolers:	N/A		

**Quality Control Preservation**

	Y	or	N	N/A
1. Trip Blank present / cooler:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
2. Trip Blank listed on COC:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
3. Samples preserved properly:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
4. VOCs headspace free:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>

**Sample Integrity - Documentation**

	Y	or	N
1. Sample labels present on bottles:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
2. Container labeling complete:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
3. Sample container label / COC agree:	<input checked="" type="checkbox"/>		<input type="checkbox"/>

**Sample Integrity - Condition**

	Y	or	N
1. Sample recvd within HT:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
2. All containers accounted for:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
3. Condition of sample:	Intact		

**Sample Integrity - Instructions**

	Y	or	N	N/A
1. Analysis requested is clear:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
2. Bottles received for unspecified tests	<input type="checkbox"/>		<input checked="" type="checkbox"/>	
3. Sufficient volume recvd for analysis:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
4. Compositing instructions clear:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
5. Filtering instructions clear:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>

Comments

# Summa Canister and Flow Controller Log

**Job Number:** JC12209  
**Account:** PROVLABR Providence Engineering  
**Project:** Valero-CAMS, Baton Rouge, LA  
**Received:** 01/08/16

32  
3

SUMMA CANISTERS													
Shipping						Receiving							
Summa ID	L	Vac " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
A1194	6	29.4	12/02/15	RC	CP8080	5W14834.D	JC12209-1	01/08/16	LM	8.5		1.3	1.52

**Accutest Bottle Order(s):**  
 VP-12/2/2015-15

**Prep Date**      **Room Temp(F)**      **Bar Pres "Hg**  
 12/02/15          70                                  29.92