

Technical Report for

Providence Engineering

Valero-CAMS, Baton Rouge, LA

Accutest Job Number: JA96838

Sampling Date: 01/07/12

Report to:

Providence Engineering


kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: 11



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Conference and/or state specific certification programs as applicable.



David N. Speis
VP, Laboratory Director

Client Service contact: Kristyn Morrison 732-329-0200

Certifications: NJ(12129), NY(10983), CA, CT, DE, FL, IL, IN, KS, KY, LA, MA, MD, MI, MT, NC, PA, RI, SC, TN, VA, WV

This report shall not be reproduced, except in its entirety, without the written approval of Accutest Laboratories.
Test results relate only to samples analyzed.

Table of Contents

-1-

Section 1: Sample Summary	3
Section 2: Sample Results	4
2.1: JA96838-1: CAMS 024	5
Section 3: Misc. Forms	7
3.1: Chain of Custody	8
3.2: Summa Canister and Flow Controller Log	11



Sample Summary

Providence Engineering

Job No: JA96838

Valero-CAMS, Baton Rouge, LA

Sample Number	Collected		Received	Matrix		Client Sample ID
	Date	Time By		Code	Type	
JA96838-1	01/07/12	13:00 KH	01/12/12	AIR	Ambient Air Grab	CAMS 024

Sample Results

Report of Analysis

Report of Analysis

Page 1 of 2

Client Sample ID:	CAMS 024	Date Sampled:	01/07/12
Lab Sample ID:	JA96838-1	Date Received:	01/12/12
Matrix:	AIR - Ambient Air Grab Summa ID: A779	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2W33915.D	1	01/18/12	YMH	n/a	n/a	V2W1428
Run #2							

Run #	Initial Volume
Run #1	400 ml
Run #2	

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	Units
67-64-1	58.08	Acetone	5.4	0.20	0.036	ppbv		13	0.48	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.024	ppbv		ND	0.44	ug/m3
71-43-2	78.11	Benzene	0.47	0.20	0.046	ppbv		1.5	0.64	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.030	ppbv		ND	1.3	ug/m3
75-25-2	252.8	Bromoform	ND	0.20	0.037	ppbv		ND	2.1	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.037	ppbv		ND	0.78	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.037	ppbv		ND	0.87	ug/m3
100-44-7	126	Benzyl Chloride	ND	0.20	0.041	ppbv		ND	1.0	ug/m3
75-15-0	76.14	Carbon disulfide	0.49	0.20	0.032	ppbv		1.5	0.62	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.027	ppbv		ND	0.92	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.039	ppbv		ND	0.53	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.028	ppbv		ND	0.98	ug/m3
74-87-3	50.49	Chloromethane	0.55	0.20	0.037	ppbv		1.1	0.41	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.041	ppbv		ND	0.63	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.031	ppbv		ND	1.0	ug/m3
56-23-5	153.8	Carbon tetrachloride	0.097	0.20	0.040	ppbv	J	0.61	1.3	ug/m3
110-82-7	84.16	Cyclohexane	0.52	0.20	0.034	ppbv		1.8	0.69	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.028	ppbv		ND	0.81	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.046	ppbv		ND	0.79	ug/m3
106-93-4	187.9	1,2-Dibromoethane	ND	0.20	0.027	ppbv		ND	1.5	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.043	ppbv		ND	0.81	ug/m3
78-87-5	113	1,2-Dichloropropane	ND	0.20	0.038	ppbv		ND	0.92	ug/m3
123-91-1	88.12	1,4-Dioxane	ND	0.20	0.056	ppbv		ND	0.72	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.53	0.20	0.038	ppbv		2.6	0.99	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.027	ppbv		ND	1.7	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.033	ppbv		ND	0.79	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.038	ppbv		ND	0.79	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.043	ppbv		ND	0.91	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.037	ppbv		ND	1.2	ug/m3
95-50-1	147	o-Dichlorobenzene	ND	0.20	0.027	ppbv		ND	1.2	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.025	ppbv		ND	1.2	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.039	ppbv		ND	0.91	ug/m3

ND = Not detected MDL - Method Detection Limit

J = Indicates an estimated value

RL = Reporting Limit

B = Indicates analyte found in associated method blank

E = Indicates value exceeds calibration range

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	CAMS 024	Date Sampled:	01/07/12
Lab Sample ID:	JA96838-1	Date Received:	01/12/12
Matrix:	AIR - Ambient Air Grab Summa ID: A779	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	Units
64-17-5	46.07	Ethanol	7.3	0.50	0.095	ppbv		14	0.94	ug/m3
100-41-4	106.2	Ethylbenzene	0.33	0.20	0.031	ppbv		1.4	0.87	ug/m3
141-78-6	88	Ethyl Acetate	2.7	0.20	0.061	ppbv		9.7	0.72	ug/m3
622-96-8	120.2	4-Ethyltoluene	0.16	0.20	0.024	ppbv	J	0.79	0.98	ug/m3
76-13-1	187.4	Freon 113	ND	0.20	0.034	ppbv		ND	1.5	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.031	ppbv		ND	1.4	ug/m3
142-82-5	100.2	Heptane	0.69	0.20	0.033	ppbv		2.8	0.82	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.046	ppbv		ND	2.1	ug/m3
110-54-3	86.17	Hexane	1.1	0.20	0.044	ppbv		3.9	0.70	ug/m3
591-78-6	100	2-Hexanone	ND	0.20	0.043	ppbv		ND	0.82	ug/m3
67-63-0	60.1	Isopropyl Alcohol	1.1	0.20	0.059	ppbv		2.7	0.49	ug/m3
75-09-2	84.94	Methylene chloride	ND	0.20	0.027	ppbv		ND	0.69	ug/m3
78-93-3	72.11	Methyl ethyl ketone	0.54	0.20	0.048	ppbv		1.6	0.59	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	0.23	0.20	0.036	ppbv		0.94	0.82	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.027	ppbv		ND	0.72	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.043	ppbv		ND	0.82	ug/m3
115-07-1	42	Propylene	ND	0.50	0.070	ppbv		ND	0.86	ug/m3
100-42-5	104.1	Styrene	0.20	0.20	0.027	ppbv		0.85	0.85	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.022	ppbv		ND	1.1	ug/m3
79-34-5	167.9	1,1,2,2-Tetrachloroethane	ND	0.20	0.030	ppbv		ND	1.4	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.030	ppbv		ND	1.1	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.051	ppbv		ND	1.5	ug/m3
95-63-6	120.2	1,2,4-Trimethylbenzene	0.58	0.20	0.024	ppbv		2.9	0.98	ug/m3
108-67-8	120.2	1,3,5-Trimethylbenzene	0.17	0.20	0.028	ppbv	J	0.84	0.98	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	0.67	0.20	0.028	ppbv		3.1	0.93	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	ND	0.20	0.032	ppbv		ND	0.61	ug/m3
127-18-4	165.8	Tetrachloroethylene	0.17	0.040	0.028	ppbv		1.2	0.27	ug/m3
109-99-9	72.11	Tetrahydrofuran	ND	0.20	0.047	ppbv		ND	0.59	ug/m3
108-88-3	92.14	Toluene	2.5	0.20	0.040	ppbv		9.4	0.75	ug/m3
79-01-6	131.4	Trichloroethylene	ND	0.040	0.033	ppbv		ND	0.21	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.27	0.20	0.042	ppbv		1.5	1.1	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.032	ppbv		ND	0.51	ug/m3
108-05-4	86	Vinyl Acetate	ND	0.20	0.057	ppbv		ND	0.70	ug/m3
	106.2	m,p-Xylene	1.0	0.20	0.031	ppbv		4.3	0.87	ug/m3
95-47-6	106.2	o-Xylene	0.39	0.20	0.031	ppbv		1.7	0.87	ug/m3
1330-20-7	106.2	Xylenes (total)	1.4	0.20	0.031	ppbv		6.1	0.87	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	98%		65-128%

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody
- Summa Canister and Flow Controller Log

CHAIN OF CUSTODY

Air Sampling Field Data Sheet

2235 US Highway 130, Dayton, NJ 08810
Tel: 732.329.0200 Fax: 732.329.3499

877045149675

Bottle Order Control #

PAGE 1 OF 1

Lab Quote #

Lab Job #

JA96838 - LA

Company Name Providence			Client/Reporting Information			Project Name Valero Refining			Weather Parameters			Requested Analysis					
Address 1201 Main St			Street			Temperature (Fahrenheit)			Start:			Maximum:					
City BR State LA Zip 70802			City Charlmette State LA			Stop:			Minimum:			TO-15					
Project Contact paul.hollis@providenceenergy.com			Project #			Atmospheric Pressure (inches of Hg)			Start:						Maximum:		
Phone # 225 766 7400 Fax # 7440			Client Purchase Order #			Stop:			Minimum:								
Sampler(s) Name(s) Karen Hudson												Other weather comment:					

Lab Sample #	Field ID / Point of Collection	Air Type				Start Sampling Information					Stop Sampling Information				
		Indoor (I) / Soil Vap (SV) / Ambient (A)	Canister Serial #	Canister Size (L or 1L)	Flow Controller Serial #	Date	Time (24 hr clock)	Canister Pressure (Psi)	Interior Temp (F)	Sampler Init.	Date	Time (24 hr clock)	Canister Pressure (Psi)	Interior Temp (F)	Sampler Init.
1	CAMS 024	A	A779	6L	-	1-6	1300	0.03	75	10H	1-7	1300	11.99	75	10H

Turnaround Time (Business Days)			Data Deliverable Information				Comments / Remarks			
Standard - 15 Days	<input checked="" type="checkbox"/>	Approved By: _____ Date: _____	All NJDEP TO-15 is mandatory Full T1				Received at Baton Rouge Service Center			
10 Day	<input type="checkbox"/>		Comm A	<input type="checkbox"/>						
5 Day	<input type="checkbox"/>		Comm B	<input type="checkbox"/>						
3 Day	<input type="checkbox"/>		Reduced T2	<input type="checkbox"/>						
2 Day	<input type="checkbox"/>		Full T1	<input type="checkbox"/>						
1 Day	<input type="checkbox"/>		Other:	<input type="checkbox"/>						
Other	<input type="checkbox"/>									

Sample Custody must be documented below each time samples change possession, including courier delivery.					
Relinquished by Laboratory:	Date Time:	Received by:	Relinquished by:	Date Time:	Received by:
1		1	[Signature]	1/10/12	2 [Signature]
Relinquished by:	Date Time:	Received by:	Relinquished by:	Date Time:	Received by:
[Signature]	1/10/12	3 FedEx	[Signature]	1/12/12	4 [Signature]
Relinquished by:	Date Time:	Received by:	Custody Seal #		
5		5			

TW

Job# JA96838
(REQUIRED)

Unused Summa Return Form

Client Conscience Office B.B. LA
Project Valero Ref.

#Summas 1 #Flow Controllers 0

Summa#'s A014 -2

Rec'd By [Signature] Rec'd Date/Time 1/12/12 08:30

Rec'd via FAX
(Attach any client paperwork, documentation, or airbills if available)

Notes

Accutest Job Number JA96838

Client:

Date / Time Received: 1/12/2012

Project:

No. Coolers: 0

Airbill #'s:

Delivery Method:

<u>Cooler Security</u>	<u>Y or N</u>		<u>Y or N</u>		
1. Custody Seals Present:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. COC Present:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2. Custody Seals Intact:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. Smpl Dates/Time OK	<input checked="" type="checkbox"/>	<input type="checkbox"/>

<u>Cooler Temperature</u>	<u>Y or N</u>	
1. Temp criteria achieved:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2. Cooler temp verification:		
3. Cooler media:		

<u>Quality Control Preservation</u>	<u>Y</u>	<u>or</u>	<u>N</u>	<u>N/A</u>
1. Trip Blank present / cooler:	<input type="checkbox"/>		<input checked="" type="checkbox"/>	<input type="checkbox"/>
2. Trip Blank listed on COC:	<input type="checkbox"/>		<input checked="" type="checkbox"/>	<input type="checkbox"/>
3. Samples preserved properly:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
4. VOCs headspace free:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>

<u>Sample Integrity - Documentation</u>	<u>Y</u>	<u>or</u>	<u>N</u>
1. Sample labels present on bottles:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
2. Container labeling complete:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
3. Sample container label / COC agree:	<input checked="" type="checkbox"/>		<input type="checkbox"/>

<u>Sample Integrity - Condition</u>	<u>Y</u>	<u>or</u>	<u>N</u>
1. Sample recvd within HT:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
2. All containers accounted for:	<input checked="" type="checkbox"/>		<input type="checkbox"/>
3. Condition of sample:			Intact

<u>Sample Integrity - Instructions</u>	<u>Y</u>	<u>or</u>	<u>N</u>	<u>N/A</u>
1. Analysis requested is clear:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
2. Bottles received for unspecified tests	<input type="checkbox"/>		<input checked="" type="checkbox"/>	
3. Sufficient volume recvd for analysis:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
4. Compositing instructions clear:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
5. Filtering instructions clear:	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>

Comments

Summa Canister and Flow Controller Log

Job Number: JA96838
Account: PROVLABR Providence Engineering
Project: Valero-CAMS, Baton Rouge, LA
Received: 01/12/12

32
3

SUMMA CANISTERS													
Shipping							Receiving						
Summa ID	L	Vac " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
A779	6	29.4	12/19/11	RC	CP5154	3W25356.D	JA96838-1	01/12/12	RC	5			1

Accutest Bottle Order(s):
 KM-12/19/2011-2

Prep Date **Room Temp(F)** **Bar Pres "Hg**
 12/19/11 70 29.92