

The results set forth herein are provided by SGS North America Inc.

e-Hardcopy 2.0
Automated Report

Technical Report for

Providence Engineering

Valero-CAMS, Baton Rouge, LA

712-001

SGS Job Number: JD40397

Sampling Date: 02/23/22

Report to:

Providence Engineering

kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: 10



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

A handwritten signature in black ink, appearing to read "Mike Earp".

Mike Earp
General Manager

Client Service contact: Jadon Schiller 732-329-0200

Certifications: NJ(12129), NY(10983), CA, CT, FL, IL, IN, KS, KY, LA, MA, MD, ME, MN, NC, OH VAP (CL0056), AK (UST-103), AZ (AZ0786), PA, RI, SC, TX, UT, VA, WV, DoD ELAP (ANAB L2248)

This report shall not be reproduced, except in its entirety, without the written approval of SGS.
Test results relate only to samples analyzed.

Table of Contents

-1-

Section 1: Sample Summary	3
Section 2: Sample Results	4
2.1: JD40397-1: CAMS 640	5
Section 3: Misc. Forms	7
3.1: Chain of Custody	8
3.2: Summa Canister and Flow Controller Log	10



Sample Summary

Providence Engineering

Job No: JD40397

Valero-CAMS, Baton Rouge, LA

Project No: 712-001

Sample Number	Collected Date	Time By	Received	Matrix Code Type	Client Sample ID
---------------	----------------	---------	----------	------------------	------------------

This report contains results reported as ND = Not detected. The following applies:

Organics ND = Not detected above the MDL

JD40397-1	02/23/22	09:45 KP	02/28/22	AIR	Ambient Air Comp.	CAMS 640
-----------	----------	----------	----------	-----	-------------------	----------

Sample Results

Report of Analysis

Report of Analysis

Client Sample ID:	CAMS 640	Date Sampled:	02/23/22
Lab Sample ID:	JD40397-1	Date Received:	02/28/22
Matrix:	AIR - Ambient Air Comp. Summa ID: A265	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	5W48393.D	1	03/13/22 01:55	TCH	n/a	n/a	V5W2000
Run #2							

Run #	Initial Volume
Run #1	400 ml
Run #2	

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
67-64-1	58.08	Acetone (2-Propanone)	3.0	0.20	0.15	ppbv		7.1	0.48	0.36	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.084	ppbv		ND	0.44	0.19	ug/m3
71-43-2	78.11	Benzene	0.28	0.20	0.062	ppbv		0.89	0.64	0.20	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.030	ppbv		ND	1.3	0.20	ug/m3
75-25-2	252.8	Bromoform	ND	0.20	0.071	ppbv		ND	2.1	0.73	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.069	ppbv		ND	0.78	0.27	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.061	ppbv		ND	0.87	0.27	ug/m3
100-44-7	126	Benzyl Chloride	ND	0.20	0.13	ppbv		ND	1.0	0.67	ug/m3
75-15-0	76.14	Carbon disulfide	ND	0.20	0.045	ppbv		ND	0.62	0.14	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.074	ppbv		ND	0.92	0.34	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.068	ppbv		ND	0.53	0.18	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.037	ppbv		ND	0.98	0.18	ug/m3
74-87-3	50.49	Chloromethane	0.71	0.20	0.090	ppbv		1.5	0.41	0.19	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.083	ppbv		ND	0.63	0.26	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.072	ppbv		ND	1.0	0.37	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	0.040	ppbv		ND	1.3	0.25	ug/m3
110-82-7	84.16	Cyclohexane	0.24	0.20	0.11	ppbv		0.83	0.69	0.38	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.057	ppbv		ND	0.81	0.23	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.059	ppbv		ND	0.79	0.23	ug/m3
106-93-4	187.9	1,2-Dibromoethane (EDB)	ND	0.20	0.097	ppbv		ND	1.5	0.75	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.070	ppbv		ND	0.81	0.28	ug/m3
78-87-5	113	1,2-Dichloropropane	ND	0.20	0.062	ppbv		ND	0.92	0.29	ug/m3
123-91-1	88.12	1,4-Dioxane	ND	0.20	0.12	ppbv		ND	0.72	0.43	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.41	0.20	0.032	ppbv		2.0	0.99	0.16	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.052	ppbv		ND	1.7	0.44	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.069	ppbv		ND	0.79	0.27	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.077	ppbv		ND	0.79	0.31	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.062	ppbv		ND	0.91	0.28	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.040	ppbv		ND	1.2	0.24	ug/m3
95-50-1	147	o-Dichlorobenzene	0.44	0.20	0.15	ppbv		2.6	1.2	0.90	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.19	ppbv		ND	1.2	1.1	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.10	ppbv		ND	0.91	0.45	ug/m3

ND = Not detected MDL = Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	CAMS 640	Date Sampled:	02/23/22
Lab Sample ID:	JD40397-1	Date Received:	02/28/22
Matrix:	AIR - Ambient Air Comp. Summa ID: A265	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
64-17-5	46.07	Ethanol	0.57	0.50	0.39	ppbv		1.1	0.94	0.73	ug/m3
100-41-4	106.2	Ethylbenzene	0.095	0.20	0.061	ppbv	J	0.41	0.87	0.26	ug/m3
141-78-6	88	Ethyl Acetate	0.44	0.20	0.10	ppbv		1.6	0.72	0.36	ug/m3
622-96-8	120.19	4-Ethyltoluene	0.11	0.20	0.095	ppbv	J	0.54	0.98	0.47	ug/m3
76-13-1	187.4	Freon 113	ND	0.20	0.031	ppbv		ND	1.5	0.24	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.050	ppbv		ND	1.4	0.35	ug/m3
142-82-5	100.2	Heptane	0.18	0.20	0.092	ppbv	J	0.74	0.82	0.38	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.062	ppbv		ND	2.1	0.66	ug/m3
110-54-3	86.18	Hexane	0.48	0.20	0.11	ppbv		1.7	0.70	0.39	ug/m3
591-78-6	100	2-Hexanone	ND	0.20	0.15	ppbv		ND	0.82	0.61	ug/m3
67-63-0	60.1	Isopropyl Alcohol	1.1	0.20	0.14	ppbv		2.7	0.49	0.34	ug/m3
75-09-2	84.94	Methylene chloride	0.17	0.20	0.056	ppbv	J	0.59	0.69	0.19	ug/m3
78-93-3	72.11	Methyl ethyl ketone	0.29	0.20	0.11	ppbv		0.86	0.59	0.32	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	ND	0.20	0.073	ppbv		ND	0.82	0.30	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.080	ppbv		ND	0.72	0.29	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.070	ppbv		ND	0.82	0.29	ug/m3
115-07-1	42	Propylene	1.1	0.50	0.14	ppbv		1.9	0.86	0.24	ug/m3
100-42-5	104.1	Styrene	ND	0.20	0.12	ppbv		ND	0.85	0.51	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.037	ppbv		ND	1.1	0.20	ug/m3
79-34-5	167.85	1,1,2,2-Tetrachloroethane	ND	0.20	0.048	ppbv		ND	1.4	0.33	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.038	ppbv		ND	1.1	0.21	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.12	ppbv		ND	1.5	0.89	ug/m3
95-63-6	120.19	1,2,4-Trimethylbenzene	0.10	0.20	0.087	ppbv	J	0.49	0.98	0.43	ug/m3
108-67-8	120.19	1,3,5-Trimethylbenzene	ND	0.20	0.080	ppbv		ND	0.98	0.39	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	0.12	0.20	0.095	ppbv	J	0.56	0.93	0.44	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	ND	0.20	0.093	ppbv		ND	0.61	0.28	ug/m3
127-18-4	165.8	Tetrachloroethylene	0.27	0.040	0.014	ppbv		1.8	0.27	0.095	ug/m3
109-99-9	72.11	Tetrahydrofuran	ND	0.20	0.090	ppbv		ND	0.59	0.27	ug/m3
108-88-3	92.14	Toluene	0.40	0.20	0.057	ppbv		1.5	0.75	0.21	ug/m3
79-01-6	131.4	Trichloroethylene	0.25	0.040	0.019	ppbv		1.3	0.21	0.10	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.21	0.20	0.036	ppbv		1.2	1.1	0.20	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.069	ppbv		ND	0.51	0.18	ug/m3
108-05-4	86	Vinyl Acetate	ND	0.20	0.11	ppbv		ND	0.70	0.39	ug/m3
	106.2	m,p-Xylene	0.40	0.20	0.14	ppbv		1.7	0.87	0.61	ug/m3
95-47-6	106.2	o-Xylene	0.17	0.20	0.077	ppbv	J	0.74	0.87	0.33	ug/m3
1330-20-7	106.2	Xylenes (total)	0.57	0.20	0.077	ppbv		2.5	0.87	0.33	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	82%		65-128%

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody
- Summa Canister and Flow Controller Log



AIR CHAIN OF CUSTODY

PAGE OF

AIR

SGS North America Inc. - Dayton
2235 Route 130, Dayton, NJ 08810
TEL: 732-329-0200 FAX: 732-329-3499
www.sgs.com/ehsusua

FED-EX Tracking # 77612966 917 Bottle Order Control # 011722-17
SGS Quote # JD40397 SGS Job # JD40397

Client / Reporting Information				Project Information				Weather Parameters				Requested Analysis				
Company Name PROVIDENCE ENGINEERING				Project Name VALERO REFINERY				Temperature (Fahrenheit)				Requested Analysis				
Address 1201 MAZON ST.				Street				Start: Maximum:		Minimum:						
City State Zip Baton Rouge LA 70802				City State MECHAN, LA				Stop: Maximum:		Minimum:						
Project Contact Drendon Kilpatrick G providenceeng.com				Project # 71K-001				Atmospheric Pressure (inches of Hg)								
Phone # 225-766-7400				Client Purchase Order #				Start: Maximum:		Minimum:						
Fax # 225-766-7440								Stop: Maximum:								
Sampler(s) Name(s) D. KILPATRICK								Other weather comment:								
Lab Sample #	Field ID / Point of Collection	Air Type	Sampling Equipment Info			Start Sampling Information					Stop Sampling Information					
			Indoor (I) Soil Vap (SV) Ambient (A)	Canister Serial #	Canister Size 6L or 1L	Flow Controller Serial #	Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.	Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.
1	CAMS 640	A	A265	6L	537	2/22	0945	30	76	TIC	2-23	0945	5.5	77	1P	Summa
Turnaround time (Business days)				Data Deliverable Information				Comments / Remarks								
<input checked="" type="checkbox"/> Standard - 15 Days <input type="checkbox"/> 10 Day <input type="checkbox"/> 5 Day <input type="checkbox"/> 3 Day <input type="checkbox"/> 2 Day <input type="checkbox"/> 1 Day <input type="checkbox"/> Other				Approved By: _____ Date: _____				All NJDEP TO-15 is mandatory Full T1 <input type="checkbox"/> Comm A <input type="checkbox"/> Comm B <input type="checkbox"/> Reduced T2 <input type="checkbox"/> Full T1 Other: _____ <input type="checkbox"/> DKQP reporting				Initial Assessment <u>GC FB</u> Label Verification _____ Sample inventory is verified upon receipt in the Laboratory				
Sample Custody must be documented below each time samples change possession, including courier delivery.																
Relinquished by Laboratory	Date Time	Received By	Relinquished by	Date Time	Received By	Relinquished by	Date Time	Received By	Relinquished by	Date Time	Received By	Relinquished by	Date Time	Received By	Relinquished by	Date Time
1 <u>[Signature]</u>	2/22 13:15	<u>[Signature]</u>	2 <u>[Signature]</u>	2/22 07:00	<u>[Signature]</u>	3 <u>[Signature]</u>	2/28/22 9:30	<u>[Signature]</u>	4 <u>[Signature]</u>			5 <u>[Signature]</u>				
Relinquished by:	Date Time:	Received By:	Relinquished by:	Date Time:	Received By:	Relinquished by:	Date Time:	Received By:	Relinquished by:	Date Time:	Received By:	Relinquished by:	Date Time:	Received By:	Relinquished by:	Date Time:
3			4			5										

Form:SM088-03D (revised 2-12-18)

http://www.sgs.com/en/terms-and-conditions

JD40397: Chain of Custody

Page 1 of 2



31
3

SGS Sample Receipt Summary

Job Number: JD40397

Client: PROVIDENCE ENG

Project: PROVIDENCE - CAMS

Date / Time Received: 2/28/2022 9:30:00 AM

Delivery Method: FEDEX

Airbill #'s:

Cooler Temps (Raw Measured) °C:

Cooler Temps (Corrected) °C:

Cooler Security

- | | | | | | |
|---------------------------|-------------------------------------|--------------------------|-----------------------|-------------------------------------|--------------------------|
| 1. Custody Seals Present: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | 3. COC Present: | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 2. Custody Seals Intact: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | 4. Smpl Dates/Time OK | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

Cooler Temperature

- | | | | |
|------------------------------|--------------------------|-----------|--------------------------|
| | <u>Y</u> | <u>or</u> | <u>N</u> |
| 1. Temp criteria achieved: | <input type="checkbox"/> | | <input type="checkbox"/> |
| 2. Cooler temp verification: | N/A | | |
| 3. Cooler media: | N/A | | |
| 4. No. Coolers: | N/A | | |

Quality Control Preservation

- | | | | | |
|---------------------------------|-------------------------------------|-----------|-------------------------------------|-------------------------------------|
| | <u>Y</u> | <u>or</u> | <u>N</u> | <u>N/A</u> |
| 1. Trip Blank present / cooler: | <input type="checkbox"/> | | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 2. Trip Blank listed on COC: | <input type="checkbox"/> | | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 3. Samples preserved properly: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> | |
| 4. VOCs headspace free: | <input type="checkbox"/> | | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

Sample Integrity - Documentation

- | | | | |
|--|-------------------------------------|-----------|--------------------------|
| | <u>Y</u> | <u>or</u> | <u>N</u> |
| 1. Sample labels present on bottles: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> |
| 2. Container labeling complete: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> |
| 3. Sample container label / COC agree: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> |

Sample Integrity - Condition

- | | | | |
|----------------------------------|-------------------------------------|-----------|--------------------------|
| | <u>Y</u> | <u>or</u> | <u>N</u> |
| 1. Sample recvd within HT: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> |
| 2. All containers accounted for: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> |
| 3. Condition of sample: | Intact | | |

Sample Integrity - Instructions

- | | | | | |
|---|-------------------------------------|-----------|-------------------------------------|-------------------------------------|
| | <u>Y</u> | <u>or</u> | <u>N</u> | <u>N/A</u> |
| 1. Analysis requested is clear: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> | |
| 2. Bottles received for unspecified tests | <input type="checkbox"/> | | <input checked="" type="checkbox"/> | |
| 3. Sufficient volume recvd for analysis: | <input checked="" type="checkbox"/> | | <input type="checkbox"/> | |
| 4. Compositing instructions clear: | <input type="checkbox"/> | | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 5. Filtering instructions clear: | <input type="checkbox"/> | | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

Test Strip Lot #s:	pH 1-12: 231619	pH 12+: 203117A	Other: (Specify)
--------------------	-----------------	-----------------	------------------

Comments

SM089-03
Rev. Date 12/7/17

JD40397: Chain of Custody

Page 2 of 2

Summa Canister and Flow Controller Log

Job Number: JD40397
Account: PROVLABR Providence Engineering
Project: Valero-CAMS, Baton Rouge, LA
Received: 02/28/22

32
3

SUMMA CANISTERS													
Shipping						Receiving							
Summa ID	Vac L	Date " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
A265	6	29.4	01/17/22	WC	CP11492	2W56998.D	JD40397-1	03/06/22	BK	6.5			1

SGS Bottle Order(s):
 JS-011722-87

Prep Date **Room Temp(F)** **Bar Pres "Hg**
 01/17/22 70 29.92