

The results set forth herein are provided by SGS North America Inc.

e-Hardcopy 2.0
Automated Report

Technical Report for

Providence Engineering

Valero-CAMS, Baton Rouge, LA

712-001

SGS Job Number: JD38294

Sampling Date: 01/12/22

Report to:

Providence Engineering

kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: 10



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

A handwritten signature in black ink, appearing to read "Mike Earp".

Mike Earp
General Manager

Client Service contact: Jadon Schiller 732-329-0200

Certifications: NJ(12129), NY(10983), CA, CT, FL, IL, IN, KS, KY, LA, MA, MD, ME, MN, NC, OH VAP (CL0056), AK (UST-103), AZ (AZ0786), PA, RI, SC, TX, UT, VA, WV, DoD ELAP (ANAB L2248)

This report shall not be reproduced, except in its entirety, without the written approval of SGS.
Test results relate only to samples analyzed.

Table of Contents

-1-

Section 1: Sample Summary	3
Section 2: Sample Results	4
2.1: JD38294-1: CAMS 633	5
Section 3: Misc. Forms	7
3.1: Chain of Custody	8
3.2: Summa Canister and Flow Controller Log	10

1

2

3



Sample Summary

Providence Engineering

Job No: JD38294

Valero-CAMS, Baton Rouge, LA
Project No: 712-001

Sample Number	Collected Date	Time By	Received	Matrix Code Type	Client Sample ID
---------------	----------------	---------	----------	------------------	------------------

This report contains results reported as ND = Not detected. The following applies:
Organics ND = Not detected above the MDL

JD38294-1	01/12/22	10:30 BK	01/17/22	AIR	Ambient Air Comp.	CAMS 633
-----------	----------	----------	----------	-----	-------------------	----------

Sample Results

Report of Analysis

Report of Analysis

Page 1 of 2

Client Sample ID:	CAMS 633	Date Sampled:	01/12/22
Lab Sample ID:	JD38294-1	Date Received:	01/17/22
Matrix:	AIR - Ambient Air Comp. Summa ID: M157	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	6W23842.D	1	01/31/22 14:54	DFT	n/a	n/a	V6W1020
Run #2							

Run #	Initial Volume
Run #1	400 ml
Run #2	

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
67-64-1	58.08	Acetone (2-Propanone)	1.4	0.20	0.15	ppbv		3.3	0.48	0.36	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.084	ppbv		ND	0.44	0.19	ug/m3
71-43-2	78.11	Benzene	0.13	0.20	0.062	ppbv	J	0.42	0.64	0.20	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.030	ppbv		ND	1.3	0.20	ug/m3
75-25-2	252.8	Bromoform	ND	0.20	0.071	ppbv		ND	2.1	0.73	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.069	ppbv		ND	0.78	0.27	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.061	ppbv		ND	0.87	0.27	ug/m3
100-44-7	126	Benzyl Chloride	ND	0.20	0.037	ppbv		ND	1.0	0.19	ug/m3
75-15-0	76.14	Carbon disulfide	ND	0.20	0.045	ppbv		ND	0.62	0.14	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.074	ppbv		ND	0.92	0.34	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.068	ppbv		ND	0.53	0.18	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.037	ppbv		ND	0.98	0.18	ug/m3
74-87-3	50.49	Chloromethane	0.48	0.20	0.090	ppbv		0.99	0.41	0.19	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.083	ppbv		ND	0.63	0.26	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.072	ppbv		ND	1.0	0.37	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	0.040	ppbv		ND	1.3	0.25	ug/m3
110-82-7	84.16	Cyclohexane	ND	0.20	0.11	ppbv		ND	0.69	0.38	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.057	ppbv		ND	0.81	0.23	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.059	ppbv		ND	0.79	0.23	ug/m3
106-93-4	187.9	1,2-Dibromoethane (EDB)	ND	0.20	0.097	ppbv		ND	1.5	0.75	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.070	ppbv		ND	0.81	0.28	ug/m3
78-87-5	113	1,2-Dichloropropane	ND	0.20	0.062	ppbv		ND	0.92	0.29	ug/m3
123-91-1	88.12	1,4-Dioxane	ND	0.20	0.12	ppbv		ND	0.72	0.43	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.40	0.20	0.032	ppbv		2.0	0.99	0.16	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.052	ppbv		ND	1.7	0.44	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.069	ppbv		ND	0.79	0.27	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.077	ppbv		ND	0.79	0.31	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.062	ppbv		ND	0.91	0.28	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.040	ppbv		ND	1.2	0.24	ug/m3
95-50-1	147	o-Dichlorobenzene	ND	0.20	0.15	ppbv		ND	1.2	0.90	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.038	ppbv		ND	1.2	0.23	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.10	ppbv		ND	0.91	0.45	ug/m3

ND = Not detected MDL = Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	CAMS 633	Date Sampled:	01/12/22
Lab Sample ID:	JD38294-1	Date Received:	01/17/22
Matrix:	AIR - Ambient Air Comp. Summa ID: M157	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
64-17-5	46.07	Ethanol	0.96	0.50	0.39	ppbv		1.8	0.94	0.73	ug/m3
100-41-4	106.2	Ethylbenzene	ND	0.20	0.061	ppbv		ND	0.87	0.26	ug/m3
141-78-6	88	Ethyl Acetate	0.14	0.20	0.10	ppbv	J	0.50	0.72	0.36	ug/m3
622-96-8	120.19	4-Ethyltoluene	ND	0.20	0.095	ppbv		ND	0.98	0.47	ug/m3
76-13-1	187.4	Freon 113	ND	0.20	0.031	ppbv		ND	1.5	0.24	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.050	ppbv		ND	1.4	0.35	ug/m3
142-82-5	100.2	Heptane	ND	0.20	0.092	ppbv		ND	0.82	0.38	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.062	ppbv		ND	2.1	0.66	ug/m3
110-54-3	86.18	Hexane	ND	0.20	0.11	ppbv		ND	0.70	0.39	ug/m3
591-78-6	100	2-Hexanone	ND	0.20	0.15	ppbv		ND	0.82	0.61	ug/m3
67-63-0	60.1	Isopropyl Alcohol	ND	0.20	0.19	ppbv		ND	0.49	0.47	ug/m3
75-09-2	84.94	Methylene chloride	ND	0.20	0.056	ppbv		ND	0.69	0.19	ug/m3
78-93-3	72.11	Methyl ethyl ketone	0.16	0.20	0.11	ppbv	J	0.47	0.59	0.32	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	ND	0.20	0.073	ppbv		ND	0.82	0.30	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.080	ppbv		ND	0.72	0.29	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.070	ppbv		ND	0.82	0.29	ug/m3
115-07-1	42	Propylene	ND	0.50	0.14	ppbv		ND	0.86	0.24	ug/m3
100-42-5	104.1	Styrene	ND	0.20	0.12	ppbv		ND	0.85	0.51	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.037	ppbv		ND	1.1	0.20	ug/m3
79-34-5	167.85	1,1,2,2-Tetrachloroethane	ND	0.20	0.048	ppbv		ND	1.4	0.33	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.038	ppbv		ND	1.1	0.21	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.12	ppbv		ND	1.5	0.89	ug/m3
95-63-6	120.19	1,2,4-Trimethylbenzene	ND	0.20	0.087	ppbv		ND	0.98	0.43	ug/m3
108-67-8	120.19	1,3,5-Trimethylbenzene	ND	0.20	0.080	ppbv		ND	0.98	0.39	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	ND	0.20	0.095	ppbv		ND	0.93	0.44	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	ND	0.20	0.093	ppbv		ND	0.61	0.28	ug/m3
127-18-4	165.8	Tetrachloroethylene	ND	0.040	0.014	ppbv		ND	0.27	0.095	ug/m3
109-99-9	72.11	Tetrahydrofuran	ND	0.20	0.090	ppbv		ND	0.59	0.27	ug/m3
108-88-3	92.14	Toluene	ND	0.20	0.057	ppbv		ND	0.75	0.21	ug/m3
79-01-6	131.4	Trichloroethylene	ND	0.040	0.019	ppbv		ND	0.21	0.10	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.21	0.20	0.036	ppbv		1.2	1.1	0.20	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.069	ppbv		ND	0.51	0.18	ug/m3
108-05-4	86	Vinyl Acetate	ND	0.20	0.11	ppbv		ND	0.70	0.39	ug/m3
	106.2	m,p-Xylene	ND	0.20	0.14	ppbv		ND	0.87	0.61	ug/m3
95-47-6	106.2	o-Xylene	ND	0.20	0.077	ppbv		ND	0.87	0.33	ug/m3
1330-20-7	106.2	Xylenes (total)	ND	0.20	0.077	ppbv		ND	0.87	0.33	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	101%		65-128%

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody
- Summa Canister and Flow Controller Log



mir

AIR CHAIN OF CUSTODY

SGS North America Inc. - Dayton
2235 Route 130, Dayton, NJ 08810
TEL 732-329-0200 FAX 732-329-3499
www.sgs.com/nehhsusa

FED-EX Tracking #	Bottle Order Control #
SGS Quote #	SGS Job #
	JD.38294

Client / Reporting Information				Project Information					Weather Parameters					Requested Analysis						
Company Name PROVENCE ENVT.				Project Name URBAN REF.					Temperature (Fahrenheit)											
Address 1201 MASON ST.				Street					Start: Maximum:											
City State Zip Dayton OH 45424				City State DAYTON OH					Stop: Minimum:											
Project Contact Brandon Patrick				Project # 712-001					Atmospheric Pressure (inches of Hg)											
E-mail brandon.patrick@providenceeng.com				Client Purchase Order #					Start: Maximum:											
Phone # 225-766-7444									Stop: Minimum:											
Fax # 225-766-7440									Other weather comment:											
Sampler(s) Name(s) Brandon Kesteven																				
Lab Sample #	Field ID / Point of Collection	Air Type Indoor (I) Soil Vap (SV) Ambient (A)	Sampling Equipment Info			Start Sampling Information					Stop Sampling Information					Requested Analysis				
			Canister Serial #	Canister Size 6L or 1L	Flow Controller Serial #	2022 Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.	2022 Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.					
1	CAMS 633	A	M157	6L	537	11/11	1030	30	64	8C	11/12	1030	55	60	R	X				
Turnaround Time (Business days)				Data Deliverable Information					Comments / Remarks											
<input checked="" type="checkbox"/> Standard - 15 Days <input type="checkbox"/> 10 Day <input type="checkbox"/> 5 Day <input type="checkbox"/> 3 Day <input type="checkbox"/> 2 Day <input type="checkbox"/> 1 Day <input type="checkbox"/> Other				Approved By: _____ Date: _____					All NJDEP TO-15 is mandatory Full T1 Comm A _____ Comm B _____ Reduced T2 _____ Full T1 _____ Other: _____ DKQP reporting _____					Sample inventory is verified upon receipt in the Laboratory						
Sample Custody must be documented below each time samples change possession, including courier delivery.																				
Relinquished by Laboratory	Date Time	Received By	Relinquished By	Date Time	Received By	Relinquished By	Date Time	Received By	Relinquished By	Date Time	Received By	Relinquished By	Date Time	Received By	Relinquished By	Date Time				
1	11/12 7:21		2	11/22 14:22																
Relinquished by:	Date Time:	Received By:	Relinquished By:	Date Time:	Received By:	Relinquished By:	Date Time:	Received By:	Relinquished By:	Date Time:	Received By:	Relinquished By:	Date Time:	Received By:	Relinquished By:	Date Time:				
5																				
Custody Seal #																				

Form:SM088-03D (revised 2-12-18)

Initial Assessment 3B E10
Label Verification

http://www.sgs.com/en/terms-and-conditions



SGS Sample Receipt Summary

Job Number: JD38294

Client: PROVIDENCE ENG

Project: PROVIDENCE - CAMS

Date / Time Received: 1/17/2022 2:02:00 PM

Delivery Method: _____

Airbill #'s: _____

Cooler Temps (Raw Measured) °C:

Cooler Temps (Corrected) °C:

Cooler Security

- | | |
|--|---|
| 1. Custody Seals Present: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | 3. COC Present: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> |
| 2. Custody Seals Intact: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | 4. Smpl Dates/Time OK: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> |

Cooler Temperature

- | | |
|--|-----|
| 1. Temp criteria achieved: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 2. Cooler temp verification: _____ | N/A |
| 3. Cooler media: _____ | N/A |
| 4. No. Coolers: _____ | N/A |

Quality Control Preservation

- | | | | | | | | | |
|---------------------------------|-------------------------------------|--------------------------|-------------------------------------|----------|--------------------------|--------------------------|-------------------------------------|------------|
| 1. Trip Blank present / cooler: | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>N/A</u> |
| 2. Trip Blank listed on COC: | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>N/A</u> |
| 3. Samples preserved properly: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <u>N/A</u> |
| 4. VOCs headspace free: | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>N/A</u> |

Sample Integrity - Documentation

- | | | | | | |
|--|-------------------------------------|--------------------------|----------|--------------------------|----------|
| 1. Sample labels present on bottles: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <u>N</u> |
| 2. Container labeling complete: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <u>N</u> |
| 3. Sample container label / COC agree: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <u>N</u> |

Sample Integrity - Condition

- | | | | | | |
|----------------------------------|-------------------------------------|--------------------------|----------|--------------------------|----------|
| 1. Sample recvd within HT: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <u>N</u> |
| 2. All containers accounted for: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <u>N</u> |
| 3. Condition of sample: | Intact | | | | |

Sample Integrity - Instructions

- | | | | | | | |
|--|-------------------------------------|-------------------------------------|----------|-------------------------------------|-------------------------------------|------------|
| 1. Analysis requested is clear: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <u>N</u> | <u>N/A</u> |
| 2. Bottles received for unspecified tests: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>Y</u> | <input checked="" type="checkbox"/> | <u>N</u> | <u>N/A</u> |
| 3. Sufficient volume recvd for analysis: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <u>N</u> | <u>N/A</u> |
| 4. Compositing instructions clear: | <input type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>N/A</u> |
| 5. Filtering instructions clear: | <input type="checkbox"/> | <input type="checkbox"/> | <u>Y</u> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <u>N/A</u> |

Test Strip Lot #s: pH 1-12: 231619 pH 12+: 203117A Other: (Specify) _____

Comments

SM089-03
Rev. Date 12/7/17

JD38294: Chain of Custody

Page 2 of 2

3.1
3

Summa Canister and Flow Controller Log

Job Number: JD38294
Account: PROVLABR Providence Engineering
Project: Valero-CAMS, Baton Rouge, LA
Received: 01/17/22

32
3

SUMMA CANISTERS													
Shipping						Receiving							
Summa ID	Vac L	Date " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
M157	6	29.4	12/31/21	MJ	CP11480	2W56848.D	JD38294-1	01/21/22	SG	6.5			1

SGS Bottle Order(s):
 JS-122721-156

Prep Date **Room Temp(F)** **Bar Pres "Hg**
 12/31/21 70 29.92