

The results set forth herein are provided by SGS North America Inc.

e-Hardcopy 2.0
Automated Report

Technical Report for

Providence Engineering

Valero-CAMS, Baton Rouge, LA

712-001

SGS Job Number: JC82115

Sampling Date: 01/20/19

Report to:

Providence Engineering

kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: 11



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

A handwritten signature in black ink, appearing to read "Brian McGuire".

Brian McGuire
General Manager

Client Service contact: Victoria Pushkova 732-329-0200

Certifications: NJ(12129), NY(10983), CA, CT, FL, IL, IN, KS, KY, LA, MA, MD, ME, MN, NC, OH VAP (CL0056), AK (UST-103), AZ (AZ0786), PA, RI, SC, TX, UT, VA, WV, DoD ELAP (ANAB L2248)

This report shall not be reproduced, except in its entirety, without the written approval of SGS.
Test results relate only to samples analyzed.

Table of Contents

-1-

| | |
|--|-----------|
| Section 1: Sample Summary | 3 |
| Section 2: Sample Results | 4 |
| 2.1: JC82115-1: CAMS 453 | 5 |
| Section 3: Misc. Forms | 8 |
| 3.1: Chain of Custody | 9 |
| 3.2: Summa Canister and Flow Controller Log | 11 |



Sample Summary

Providence Engineering

Job No: JC82115

Valero-CAMS, Baton Rouge, LA
Project No: 712-001

| Sample Number | Collected | | Matrix | | | Client Sample ID |
|---------------|-----------|-----------|----------|------|------------------|------------------|
| | Date | Time By | Received | Code | Type | |
| JC82115-1 | 01/20/19 | 09:20 JMK | 01/31/19 | AIR | Ambient Air Grab | CAMS 453 |

Sample Results

Report of Analysis

Report of Analysis

| | | | |
|-------------------|---------------------------------------|-----------------|----------|
| Client Sample ID: | CAMS 453 | Date Sampled: | 01/20/19 |
| Lab Sample ID: | JC82115-1 | Date Received: | 01/31/19 |
| Matrix: | AIR - Ambient Air Grab Summa ID: A336 | Percent Solids: | n/a |
| Method: | TO-15 | | |
| Project: | Valero-CAMS, Baton Rouge, LA | | |

| Run # | File ID | DF | Analyzed | By | Prep Date | Prep Batch | Analytical Batch |
|--------|-----------|----|----------------|-----|-----------|------------|------------------|
| Run #1 | 6W10460.D | 1 | 02/05/19 14:47 | TCH | n/a | n/a | V6W413 |
| Run #2 | | | | | | | |

| Run # | Initial Volume |
|--------|----------------|
| Run #1 | 400 ml |
| Run #2 | |

VOA TO15 List

| CAS No. | MW | Compound | Result | RL | MDL | Units | Q | Result | RL | MDL | Units |
|------------|-------|------------------------------|--------|------|--------|-------|---|--------|------|-------|-------|
| 67-64-1 | 58.08 | Acetone | 2.8 | 0.20 | 0.11 | ppbv | | 6.7 | 0.48 | 0.26 | ug/m3 |
| 106-99-0 | 54.09 | 1,3-Butadiene | ND | 0.20 | 0.046 | ppbv | | ND | 0.44 | 0.10 | ug/m3 |
| 71-43-2 | 78.11 | Benzene | ND | 0.20 | 0.012 | ppbv | | ND | 0.64 | 0.038 | ug/m3 |
| 75-27-4 | 163.8 | Bromodichloromethane | ND | 0.20 | 0.027 | ppbv | | ND | 1.3 | 0.18 | ug/m3 |
| 75-25-2 | 252.8 | Bromoform | ND | 0.20 | 0.037 | ppbv | | ND | 2.1 | 0.38 | ug/m3 |
| 74-83-9 | 94.94 | Bromomethane | ND | 0.20 | 0.022 | ppbv | | ND | 0.78 | 0.085 | ug/m3 |
| 593-60-2 | 106.9 | Bromoethene | ND | 0.20 | 0.022 | ppbv | | ND | 0.87 | 0.096 | ug/m3 |
| 100-44-7 | 126 | Benzyl Chloride ^a | ND | 0.20 | 0.057 | ppbv | | ND | 1.0 | 0.29 | ug/m3 |
| 75-15-0 | 76.14 | Carbon disulfide | ND | 0.20 | 0.024 | ppbv | | ND | 0.62 | 0.075 | ug/m3 |
| 108-90-7 | 112.6 | Chlorobenzene | ND | 0.20 | 0.026 | ppbv | | ND | 0.92 | 0.12 | ug/m3 |
| 75-00-3 | 64.52 | Chloroethane | ND | 0.20 | 0.048 | ppbv | | ND | 0.53 | 0.13 | ug/m3 |
| 67-66-3 | 119.4 | Chloroform | ND | 0.20 | 0.020 | ppbv | | ND | 0.98 | 0.098 | ug/m3 |
| 74-87-3 | 50.49 | Chloromethane | 0.59 | 0.20 | 0.015 | ppbv | | 1.2 | 0.41 | 0.031 | ug/m3 |
| 107-05-1 | 76.53 | 3-Chloropropene | ND | 0.20 | 0.040 | ppbv | | ND | 0.63 | 0.13 | ug/m3 |
| 95-49-8 | 126.6 | 2-Chlorotoluene | ND | 0.20 | 0.025 | ppbv | | ND | 1.0 | 0.13 | ug/m3 |
| 56-23-5 | 153.8 | Carbon tetrachloride | ND | 0.20 | 0.024 | ppbv | | ND | 1.3 | 0.15 | ug/m3 |
| 110-82-7 | 84.16 | Cyclohexane | ND | 0.20 | 0.022 | ppbv | | ND | 0.69 | 0.076 | ug/m3 |
| 75-34-3 | 98.96 | 1,1-Dichloroethane | ND | 0.20 | 0.012 | ppbv | | ND | 0.81 | 0.049 | ug/m3 |
| 75-35-4 | 96.94 | 1,1-Dichloroethylene | ND | 0.20 | 0.017 | ppbv | | ND | 0.79 | 0.067 | ug/m3 |
| 106-93-4 | 187.9 | 1,2-Dibromoethane | ND | 0.20 | 0.018 | ppbv | | ND | 1.5 | 0.14 | ug/m3 |
| 107-06-2 | 98.96 | 1,2-Dichloroethane | ND | 0.20 | 0.021 | ppbv | | ND | 0.81 | 0.085 | ug/m3 |
| 78-87-5 | 113 | 1,2-Dichloropropane | ND | 0.20 | 0.019 | ppbv | | ND | 0.92 | 0.088 | ug/m3 |
| 123-91-1 | 88.12 | 1,4-Dioxane | ND | 0.20 | 0.052 | ppbv | | ND | 0.72 | 0.19 | ug/m3 |
| 75-71-8 | 120.9 | Dichlorodifluoromethane | 0.32 | 0.20 | 0.017 | ppbv | | 1.6 | 0.99 | 0.084 | ug/m3 |
| 124-48-1 | 208.3 | Dibromochloromethane | ND | 0.20 | 0.033 | ppbv | | ND | 1.7 | 0.28 | ug/m3 |
| 156-60-5 | 96.94 | trans-1,2-Dichloroethylene | ND | 0.20 | 0.0073 | ppbv | | ND | 0.79 | 0.029 | ug/m3 |
| 156-59-2 | 96.94 | cis-1,2-Dichloroethylene | ND | 0.20 | 0.012 | ppbv | | ND | 0.79 | 0.048 | ug/m3 |
| 10061-01-5 | 111 | cis-1,3-Dichloropropene | ND | 0.20 | 0.020 | ppbv | | ND | 0.91 | 0.091 | ug/m3 |
| 541-73-1 | 147 | m-Dichlorobenzene | ND | 0.20 | 0.019 | ppbv | | ND | 1.2 | 0.11 | ug/m3 |
| 95-50-1 | 147 | o-Dichlorobenzene | ND | 0.20 | 0.022 | ppbv | | ND | 1.2 | 0.13 | ug/m3 |
| 106-46-7 | 147 | p-Dichlorobenzene | ND | 0.20 | 0.018 | ppbv | | ND | 1.2 | 0.11 | ug/m3 |
| 10061-02-6 | 111 | trans-1,3-Dichloropropene | ND | 0.20 | 0.020 | ppbv | | ND | 0.91 | 0.091 | ug/m3 |

ND = Not detected MDL = Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

| | | | |
|--------------------------|------------------------------|------------------------|----------|
| Client Sample ID: | CAMS 453 | Date Sampled: | 01/20/19 |
| Lab Sample ID: | JC82115-1 | Date Received: | 01/31/19 |
| Matrix: | AIR - Ambient Air Grab | Summa ID: | A336 |
| Method: | TO-15 | Percent Solids: | n/a |
| Project: | Valero-CAMS, Baton Rouge, LA | | |

VOA TO15 List

| CAS No. | MW | Compound | Result | RL | MDL | Units | Q | Result | RL | MDL | Units |
|-----------|--------|---------------------------|--------|-------|-------|-------|---|--------|------|-------|-------|
| 64-17-5 | 46.07 | Ethanol | 17.7 | 0.50 | 0.22 | ppbv | | 33.4 | 0.94 | 0.41 | ug/m3 |
| 100-41-4 | 106.2 | Ethylbenzene | ND | 0.20 | 0.015 | ppbv | | ND | 0.87 | 0.065 | ug/m3 |
| 141-78-6 | 88 | Ethyl Acetate | ND | 0.20 | 0.038 | ppbv | | ND | 0.72 | 0.14 | ug/m3 |
| 622-96-8 | 120.2 | 4-Ethyltoluene | ND | 0.20 | 0.030 | ppbv | | ND | 0.98 | 0.15 | ug/m3 |
| 76-13-1 | 187.4 | Freon 113 | ND | 0.20 | 0.017 | ppbv | | ND | 1.5 | 0.13 | ug/m3 |
| 76-14-2 | 170.9 | Freon 114 | ND | 0.20 | 0.019 | ppbv | | ND | 1.4 | 0.13 | ug/m3 |
| 142-82-5 | 100.2 | Heptane | ND | 0.20 | 0.018 | ppbv | | ND | 0.82 | 0.074 | ug/m3 |
| 87-68-3 | 260.8 | Hexachlorobutadiene | ND | 0.20 | 0.046 | ppbv | | ND | 2.1 | 0.49 | ug/m3 |
| 110-54-3 | 86.17 | Hexane | ND | 0.20 | 0.011 | ppbv | | ND | 0.70 | 0.039 | ug/m3 |
| 591-78-6 | 100 | 2-Hexanone | ND | 0.20 | 0.036 | ppbv | | ND | 0.82 | 0.15 | ug/m3 |
| 67-63-0 | 60.1 | Isopropyl Alcohol | ND | 0.20 | 0.065 | ppbv | | ND | 0.49 | 0.16 | ug/m3 |
| 75-09-2 | 84.94 | Methylene chloride | ND | 0.20 | 0.015 | ppbv | | ND | 0.69 | 0.052 | ug/m3 |
| 78-93-3 | 72.11 | Methyl ethyl ketone | ND | 0.20 | 0.042 | ppbv | | ND | 0.59 | 0.12 | ug/m3 |
| 108-10-1 | 100.2 | Methyl Isobutyl Ketone | ND | 0.20 | 0.036 | ppbv | | ND | 0.82 | 0.15 | ug/m3 |
| 1634-04-4 | 88.15 | Methyl Tert Butyl Ether | ND | 0.20 | 0.019 | ppbv | | ND | 0.72 | 0.069 | ug/m3 |
| 80-62-6 | 100.12 | Methylmethacrylate | ND | 0.20 | 0.033 | ppbv | | ND | 0.82 | 0.14 | ug/m3 |
| 115-07-1 | 42 | Propylene | ND | 0.50 | 0.016 | ppbv | | ND | 0.86 | 0.027 | ug/m3 |
| 100-42-5 | 104.1 | Styrene | ND | 0.20 | 0.019 | ppbv | | ND | 0.85 | 0.081 | ug/m3 |
| 71-55-6 | 133.4 | 1,1,1-Trichloroethane | ND | 0.20 | 0.033 | ppbv | | ND | 1.1 | 0.18 | ug/m3 |
| 79-34-5 | 167.9 | 1,1,2,2-Tetrachloroethane | ND | 0.20 | 0.027 | ppbv | | ND | 1.4 | 0.19 | ug/m3 |
| 79-00-5 | 133.4 | 1,1,2-Trichloroethane | ND | 0.20 | 0.030 | ppbv | | ND | 1.1 | 0.16 | ug/m3 |
| 120-82-1 | 181.5 | 1,2,4-Trichlorobenzene | ND | 0.20 | 0.089 | ppbv | | ND | 1.5 | 0.66 | ug/m3 |
| 95-63-6 | 120.2 | 1,2,4-Trimethylbenzene | ND | 0.20 | 0.033 | ppbv | | ND | 0.98 | 0.16 | ug/m3 |
| 108-67-8 | 120.2 | 1,3,5-Trimethylbenzene | ND | 0.20 | 0.034 | ppbv | | ND | 0.98 | 0.17 | ug/m3 |
| 540-84-1 | 114.2 | 2,2,4-Trimethylpentane | ND | 0.20 | 0.022 | ppbv | | ND | 0.93 | 0.10 | ug/m3 |
| 75-65-0 | 74.12 | Tertiary Butyl Alcohol | ND | 0.20 | 0.014 | ppbv | | ND | 0.61 | 0.042 | ug/m3 |
| 127-18-4 | 165.8 | Tetrachloroethylene | ND | 0.040 | 0.031 | ppbv | | ND | 0.27 | 0.21 | ug/m3 |
| 109-99-9 | 72.11 | Tetrahydrofuran | ND | 0.20 | 0.050 | ppbv | | ND | 0.59 | 0.15 | ug/m3 |
| 108-88-3 | 92.14 | Toluene | ND | 0.20 | 0.014 | ppbv | | ND | 0.75 | 0.053 | ug/m3 |
| 79-01-6 | 131.4 | Trichloroethylene | ND | 0.040 | 0.019 | ppbv | | ND | 0.21 | 0.10 | ug/m3 |
| 75-69-4 | 137.4 | Trichlorofluoromethane | ND | 0.20 | 0.028 | ppbv | | ND | 1.1 | 0.16 | ug/m3 |
| 75-01-4 | 62.5 | Vinyl chloride | ND | 0.20 | 0.022 | ppbv | | ND | 0.51 | 0.056 | ug/m3 |
| 108-05-4 | 86 | Vinyl Acetate | ND | 0.20 | 0.034 | ppbv | | ND | 0.70 | 0.12 | ug/m3 |
| | 106.2 | m,p-Xylene | ND | 0.20 | 0.034 | ppbv | | ND | 0.87 | 0.15 | ug/m3 |
| 95-47-6 | 106.2 | o-Xylene | ND | 0.20 | 0.017 | ppbv | | ND | 0.87 | 0.074 | ug/m3 |
| 1330-20-7 | 106.2 | Xylenes (total) | ND | 0.20 | 0.017 | ppbv | | ND | 0.87 | 0.074 | ug/m3 |

| CAS No. | Surrogate Recoveries | Run# 1 | Run# 2 | Limits |
|----------|----------------------|--------|--------|---------|
| 460-00-4 | 4-Bromofluorobenzene | 98% | | 65-128% |

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

| | | |
|--|-----------------------|--------------------------------|
| Client Sample ID: CAMS 453 | | |
| Lab Sample ID: JC82115-1 | | Date Sampled: 01/20/19 |
| Matrix: AIR - Ambient Air Grab | Summa ID: A336 | Date Received: 01/31/19 |
| Method: TO-15 | | Percent Solids: n/a |
| Project: Valero-CAMS, Baton Rouge, LA | | |

VOA TO15 List

| CAS No. | MW | Compound | Result | RL | MDL | Units | Q | Result | RL | MDL | Units |
|---------|----|----------|--------|----|-----|-------|---|--------|----|-----|-------|
|---------|----|----------|--------|----|-----|-------|---|--------|----|-----|-------|

(a) Associated CCV outside of control limits high, sample was ND.

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody
- Summa Canister and Flow Controller Log

SGS North America Inc. - Dayton
2235 Route 130, Dayton, NJ 08810
TEL: 732-329-0200 FAX: 732-329-3499
www.sgs.com/ehsusa

FED-EX Tracking #
SGS Quote #
Request Under Control #
SGS Job #
VP 811419-94
JC82115

| Client / Reporting Information | | | Project Information | | | Weather Parameters | | | Requested Analysis | | |
|---|--|--|------------------------------|--|--|-------------------------------------|--|--|------------------------|--|--|
| Company Name: PROVIDENCE | | | Project Name: VALECO MECANIX | | | Temperature (Fahrenheit) | | | | | |
| Address: 1201 MAIN | | | Street: | | | Start: Maximum: | | | | | |
| City: BATON ROUGE LA 70802 | | | City: State: | | | Stop: Minimum: | | | | | |
| Project Contact: PAUL HOLLIS PAULHOLLIS@PROVIDENCEENG.COM | | | Project #: 0712-001 | | | Atmospheric Pressure (inches of Hg) | | | | | |
| Phone #: 225-766-7440 Fax #: 225-766-7440 | | | Client Purchase Order #: | | | Start: Maximum: | | | | | |
| Sampler(s) Name(s): | | | | | | Stop: Minimum: | | | Other weather comment: | | |

| Lab Sample # | Field ID / Point of Collection | Air Type | | Sampling Equipment Info | | | Start Sampling Information | | | | | Stop Sampling Information | | | | |
|--------------|--------------------------------|--|-------------------|---------------------------|-----------------------------|---------|----------------------------|----------------------------|----------------------|---------------|---------|---------------------------|----------------------------|----------------------|---------------|--|
| | | Indoor (I) Soil Vap (SV) Ambient (A) | Canister Serial # | Canister Size 6L or 1L | Flow Controller Serial # | Date | Time (24hr clock) | Canister Pressure ("Hg) | Interior Temp (F) | Sampler Init. | Date | Time (24hr clock) | Canister Pressure ("Hg) | Interior Temp (F) | Sampler Init. | |
| 1 | CAMS 453 | A | A336 | 6L | 537 | 1/19/19 | 0920 | 30 | 70 | PAN | 1/20/19 | 0920 | 5 | 50 | SMK | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | |

70-15

Summary

| Turnaround Time (Business days) | | Data Deliverable Information | | Comments / Remarks | |
|--|--------------------|---|--|---|--|
| <input checked="" type="checkbox"/> Standard - 15 Days | Approved By: _____ | All NJDEP TO-15 is mandatory Full T1 | | INITIAL ASESSMENT <u>PAH</u> | |
| <input type="checkbox"/> 10 Day | Date: _____ | Comm A | | | |
| <input type="checkbox"/> 5 Day | | <input type="checkbox"/> Reduced T2 | | | |
| <input type="checkbox"/> 3 Day | | <input type="checkbox"/> Full T1 | | | |
| <input type="checkbox"/> 2 Day | | <input type="checkbox"/> Other: _____ | | | |
| <input type="checkbox"/> 1 Day | | <input type="checkbox"/> DKQP reporting | | Sample inventory is turned over to client in the laboratory | |
| <input type="checkbox"/> Other | | | | LABEL VERIFICATION _____ | |

| | | | | | |
|--|----------------------------|------------------------------------|--|-----------------------------|----------------------|
| Sample Custody must be documented below each time samples change possession, including courier delivery. | | | | | |
| Relinquished by Laboratory: 1 P. Shah | Date Time: 1/15/19 7:15 | Received By: <i>[Signature]</i> | Relinquished By: <i>[Signature]</i> | Date Time: 1/16/19 12:00 | Received By: 2 FX |
| Relinquished by: 3 FX | Date Time: 1/21/19 9:30 | Received By: <i>[Signature]</i> | Relinquished By: <i>[Signature]</i> | Date Time: | Received By: 4 |
| Relinquished by: | Date Time: | Received By: | Custody Seal # | | |
| 5 | | | | | |

31
3

SGS Sample Receipt Summary

Job Number: JC82115

Client: PROVIDENCE ENG

Project: PROVIDENCE - CAMS

Date / Time Received: 1/31/2019 9:30:00 AM

Delivery Method: _____

Airbill #'s: _____

Cooler Temps (Raw Measured) °C:

Cooler Temps (Corrected) °C:

Cooler Security

- | | |
|--|---|
| 1. Custody Seals Present: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | 3. COC Present: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> |
| 2. Custody Seals Intact: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | 4. Smpl Dates/Time OK: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> |

Cooler Temperature

- | | |
|--|-----|
| 1. Temp criteria achieved: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 2. Cooler temp verification: _____ | N/A |
| 3. Cooler media: _____ | N/A |
| 4. No. Coolers: _____ | N/A |

Quality Control Preservation

- | | | | |
|---------------------------------|--|--|--|
| 1. Trip Blank present / cooler: | <input type="checkbox"/> <u>Y</u> | <input checked="" type="checkbox"/> <u>N</u> | <input type="checkbox"/> <u>N/A</u> |
| 2. Trip Blank listed on COC: | <input type="checkbox"/> <u>Y</u> | <input checked="" type="checkbox"/> <u>N</u> | <input type="checkbox"/> <u>N/A</u> |
| 3. Samples preserved properly: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> | <input type="checkbox"/> <u>N/A</u> |
| 4. VOCs headspace free: | <input type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> | <input checked="" type="checkbox"/> <u>N/A</u> |

Sample Integrity - Documentation

- | | | |
|--|--|-----------------------------------|
| 1. Sample labels present on bottles: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> |
| 2. Container labeling complete: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> |
| 3. Sample container label / COC agree: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> |

Sample Integrity - Condition

- | | | |
|----------------------------------|--|-----------------------------------|
| 1. Sample recvd within HT: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> |
| 2. All containers accounted for: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> |
| 3. Condition of sample: | Intact | |

Sample Integrity - Instructions

- | | | | | |
|--|--|--|--|--|
| 1. Analysis requested is clear: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> | | |
| 2. Bottles received for unspecified tests: | <input type="checkbox"/> <u>Y</u> | <input checked="" type="checkbox"/> <u>N</u> | | |
| 3. Sufficient volume recvd for analysis: | <input checked="" type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> | | |
| 4. Compositing instructions clear: | <input type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> | <input checked="" type="checkbox"/> <u>N/A</u> | |
| 5. Filtering instructions clear: | <input type="checkbox"/> <u>Y</u> | <input type="checkbox"/> <u>N</u> | <input checked="" type="checkbox"/> <u>N/A</u> | |

| | | | |
|--------------------|-----------------|----------------|------------------------|
| Test Strip Lot #s: | pH 1-12: 206717 | pH 12+: 208717 | Other: (Specify) _____ |
|--------------------|-----------------|----------------|------------------------|

Comments

SM089-03
Rev. Date 12/7/17

JC82115: Chain of Custody

Page 2 of 2

Summa Canister and Flow Controller Log

Job Number: JC82115
Account: PROVLABR Providence Engineering
Project: Valero-CAMS, Baton Rouge, LA
Received: 01/31/19

32
3

| SUMMA CANISTERS | | | | | | | | | | | | | |
|-----------------|-------|-----------|----------|----|-----------|------------|---------------|----------|----|----------|-----------|------------|----------|
| Shipping | | | | | | | Receiving | | | | | | |
| Summa ID | Vac L | Date " Hg | Date Out | By | SCC Batch | SCC FileID | Sample Number | Date In | By | Vac " Hg | Pres psig | Final psig | Dil Fact |
| A336 | 6 | 29.4 | 01/15/19 | PC | CP10146 | 5W34657.D | JC82115-1 | 02/01/19 | PC | 5.5 | | | 1 |

SGS Bottle Order(s):
 VP-011419-94

Prep Date **Room Temp(F)** **Bar Pres "Hg**
 01/15/19 70 29.92