

The results set forth herein are provided by SGS North America Inc.

e-Hardcopy 2.0
Automated Report

Technical Report for

Providence Engineering

Valero-CAMS, Baton Rouge, LA

712-001

SGS Job Number: JD23480

Sampling Date: 04/09/21

Report to:

Providence Engineering

kevincalhoun@providenceeng.com

ATTN: Kevin Calhoun

Total number of pages in report: 11



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

A handwritten signature in black ink that reads "Caitlin Brice".

Caitlin Brice, M.S.
General Manager

Client Service contact: Shalini Williams 732-329-0200

Certifications: NJ(12129), NY(10983), CA, CT, FL, IL, IN, KS, KY, LA, MA, MD, ME, MN, NC, OH VAP (CL0056), AK (UST-103), AZ (AZ0786), PA, RI, SC, TX, UT, VA, WV, DoD ELAP (ANAB L2248)

This report shall not be reproduced, except in its entirety, without the written approval of SGS.
Test results relate only to samples analyzed.

Table of Contents

-1-

Section 1: Sample Summary	3
Section 2: Sample Results	4
2.1: JD23480-1: CAMS 444	5
Section 3: Misc. Forms	8
3.1: Chain of Custody	9
3.2: Summa Canister and Flow Controller Log	11



Sample Summary

Providence Engineering

Job No: JD23480

Valero-CAMS, Baton Rouge, LA

Project No: 712-001

Sample Number	Collected Date	Time By	Received	Matrix Code Type	Client Sample ID
---------------	----------------	---------	----------	------------------	------------------

This report contains results reported as ND = Not detected. The following applies:

Organics ND = Not detected above the MDL

JD23480-1	04/09/21	12:00 DJ	04/15/21	AIR Ambient Air Comp.	CAMS 444
-----------	----------	----------	----------	-----------------------	----------

Sample Results

Report of Analysis

Report of Analysis

Client Sample ID:	CAMS 444	Date Sampled:	04/09/21
Lab Sample ID:	JD23480-1	Date Received:	04/15/21
Matrix:	AIR - Ambient Air Comp. Summa ID: M009	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2W53290.D	1	04/23/21 17:45	TCH	n/a	n/a	V2W2359
Run #2							

Run #	Initial Volume
Run #1	400 ml
Run #2	

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
67-64-1	58.08	Acetone	5.5	0.20	0.11	ppbv		13	0.48	0.26	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.046	ppbv		ND	0.44	0.10	ug/m3
71-43-2	78.11	Benzene	0.31	0.20	0.012	ppbv		0.99	0.64	0.038	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.027	ppbv		ND	1.3	0.18	ug/m3
75-25-2	252.8	Bromoform ^a	ND	0.20	0.037	ppbv		ND	2.1	0.38	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.022	ppbv		ND	0.78	0.085	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.022	ppbv		ND	0.87	0.096	ug/m3
100-44-7	126	Benzyl Chloride ^a	ND	0.20	0.057	ppbv		ND	1.0	0.29	ug/m3
75-15-0	76.14	Carbon disulfide	ND	0.20	0.024	ppbv		ND	0.62	0.075	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.026	ppbv		ND	0.92	0.12	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.048	ppbv		ND	0.53	0.13	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.020	ppbv		ND	0.98	0.098	ug/m3
74-87-3	50.49	Chloromethane	1.1	0.20	0.015	ppbv		2.3	0.41	0.031	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.040	ppbv		ND	0.63	0.13	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.025	ppbv		ND	1.0	0.13	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	0.024	ppbv		ND	1.3	0.15	ug/m3
110-82-7	84.16	Cyclohexane	0.55	0.20	0.022	ppbv		1.9	0.69	0.076	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.012	ppbv		ND	0.81	0.049	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.017	ppbv		ND	0.79	0.067	ug/m3
106-93-4	187.9	1,2-Dibromoethane	ND	0.20	0.018	ppbv		ND	1.5	0.14	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.021	ppbv		ND	0.81	0.085	ug/m3
78-87-5	113	1,2-Dichloropropane	ND	0.20	0.019	ppbv		ND	0.92	0.088	ug/m3
123-91-1	88.12	1,4-Dioxane	ND	0.20	0.052	ppbv		ND	0.72	0.19	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.35	0.20	0.017	ppbv		1.7	0.99	0.084	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.033	ppbv		ND	1.7	0.28	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.0073	ppbv		ND	0.79	0.029	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.012	ppbv		ND	0.79	0.048	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.020	ppbv		ND	0.91	0.091	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.019	ppbv		ND	1.2	0.11	ug/m3
95-50-1	147	o-Dichlorobenzene	ND	0.20	0.022	ppbv		ND	1.2	0.13	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.018	ppbv		ND	1.2	0.11	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.020	ppbv		ND	0.91	0.091	ug/m3

ND = Not detected MDL = Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	CAMS 444	Date Sampled:	04/09/21
Lab Sample ID:	JD23480-1	Date Received:	04/15/21
Matrix:	AIR - Ambient Air Comp. Summa ID: M009	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero-CAMS, Baton Rouge, LA		

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
64-17-5	46.07	Ethanol	1.6	0.50	0.22	ppbv		3.0	0.94	0.41	ug/m3
100-41-4	106.2	Ethylbenzene	0.11	0.20	0.015	ppbv	J	0.48	0.87	0.065	ug/m3
141-78-6	88	Ethyl Acetate	0.42	0.20	0.038	ppbv		1.5	0.72	0.14	ug/m3
622-96-8	120.2	4-Ethyltoluene	ND	0.20	0.030	ppbv		ND	0.98	0.15	ug/m3
76-13-1	187.4	Freon 113	ND	0.20	0.017	ppbv		ND	1.5	0.13	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.019	ppbv		ND	1.4	0.13	ug/m3
142-82-5	100.2	Heptane	0.71	0.20	0.018	ppbv		2.9	0.82	0.074	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.046	ppbv		ND	2.1	0.49	ug/m3
110-54-3	86.17	Hexane	1.7	0.20	0.011	ppbv		6.0	0.70	0.039	ug/m3
591-78-6	100	2-Hexanone	0.12	0.20	0.036	ppbv	J	0.49	0.82	0.15	ug/m3
67-63-0	60.1	Isopropyl Alcohol	0.56	0.20	0.065	ppbv		1.4	0.49	0.16	ug/m3
75-09-2	84.94	Methylene chloride	ND	0.20	0.015	ppbv		ND	0.69	0.052	ug/m3
78-93-3	72.11	Methyl ethyl ketone	0.81	0.20	0.042	ppbv		2.4	0.59	0.12	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	ND	0.20	0.036	ppbv		ND	0.82	0.15	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.019	ppbv		ND	0.72	0.069	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.033	ppbv		ND	0.82	0.14	ug/m3
115-07-1	42	Propylene	0.53	0.50	0.016	ppbv		0.91	0.86	0.027	ug/m3
100-42-5	104.1	Styrene	0.10	0.20	0.019	ppbv	J	0.43	0.85	0.081	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.033	ppbv		ND	1.1	0.18	ug/m3
79-34-5	167.9	1,1,2,2-Tetrachloroethane	ND	0.20	0.027	ppbv		ND	1.4	0.19	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.030	ppbv		ND	1.1	0.16	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.089	ppbv		ND	1.5	0.66	ug/m3
95-63-6	120.2	1,2,4-Trimethylbenzene	0.12	0.20	0.033	ppbv	J	0.59	0.98	0.16	ug/m3
108-67-8	120.2	1,3,5-Trimethylbenzene	ND	0.20	0.034	ppbv		ND	0.98	0.17	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	0.18	0.20	0.022	ppbv	J	0.84	0.93	0.10	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	ND	0.20	0.014	ppbv		ND	0.61	0.042	ug/m3
127-18-4	165.8	Tetrachloroethylene	4.6	0.040	0.031	ppbv		31	0.27	0.21	ug/m3
109-99-9	72.11	Tetrahydrofuran	ND	0.20	0.050	ppbv		ND	0.59	0.15	ug/m3
108-88-3	92.14	Toluene	0.60	0.20	0.014	ppbv		2.3	0.75	0.053	ug/m3
79-01-6	131.4	Trichloroethylene	0.72	0.040	0.019	ppbv		3.9	0.21	0.10	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.20	0.20	0.028	ppbv		1.1	1.1	0.16	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.022	ppbv		ND	0.51	0.056	ug/m3
108-05-4	86	Vinyl Acetate	ND	0.20	0.034	ppbv		ND	0.70	0.12	ug/m3
	106.2	m,p-Xylene	0.31	0.20	0.034	ppbv		1.3	0.87	0.15	ug/m3
95-47-6	106.2	o-Xylene	0.12	0.20	0.017	ppbv	J	0.52	0.87	0.074	ug/m3
1330-20-7	106.2	Xylenes (total)	0.43	0.20	0.017	ppbv		1.9	0.87	0.074	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	103%		65-128%

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID: CAMS 444		
Lab Sample ID: JD23480-1		Date Sampled: 04/09/21
Matrix: AIR - Ambient Air Comp.	Summa ID: M009	Date Received: 04/15/21
Method: TO-15		Percent Solids: n/a
Project: Valero-CAMS, Baton Rouge, LA		

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
---------	----	----------	--------	----	-----	-------	---	--------	----	-----	-------

(a) Associated CCV outside of control limits high, sample was ND.

ND = Not detected	MDL = Method Detection Limit	J = Indicates an estimated value
RL = Reporting Limit		B = Indicates analyte found in associated method blank
E = Indicates value exceeds calibration range		N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody
- Summa Canister and Flow Controller Log

AIR

CHAIN OF CUSTODY

SGS North America Inc. - Houston Air Sampling Field Data Sheet

JD 23480



10165 Harwin Dr, Ste 150 Houston, TX 77036
TEL: 713-271-4700 FAX: 713-271-4770
www.sgs.com/ehsusa

FED-EX Tracking # 2341369605
Lab Quote #

Bottle Order Control #
Lab Job #

PAGE 1 OF 1

Client/Reporting Information				Project Information				Weather Parameters					Requested Analysis			
Company Name Providence Engineering				Project Name Valero Refining				Temperature (Fahrenheit)								
Address 1701 main St.				Street				Start:		Maximum:						
City Baton Rouge				City Meroux				Stop:		Minimum:						
Project Contact Brandon K. Patrick				Project #				Atmospheric Pressure (inches of Hg)								
E-mail Brandon.K.Patrick@Providenceeng.com				Client Purchase Order #				Start:		Maximum:						
Phone # 225-766-7700								Stop:		Minimum:						
Sampler(s) Name(s) Derek Jarman								Other weather comment:								
Lab Sample #	Field ID / Point of Collection	Air Type		Sampling Equipment Info			Start Sampling Information					Stop Sampling Information				
		Indoor(I) Soil Vap(SV) Ambient(A) Grab (G)	Canister Serial # Tedlar	Size 6L or 1L	Flow Controller Serial #	Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.	Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.	
1	Cams 444	A	m009	6L	59205	4-9-21	1200	30	77	05	4-9-21	1200	5.5	75	05	X
Turnaround Time		Approved By: _____		Date: _____		Data Deliverable Information					Comments / Remarks					
Standard		All NJDEP TO-15 is mandatory Full T1		Comm A					Comm B							
10 Business Days		Reduced T2		Full T1					Other:							
5 Business Days																
3 Business Days																
2 Business Days																
1 Business Day																
Other																
INITIAL ASSESSMENT <u>3BDM</u>																
LABEL VERIFICATION																
Sample Custody must be documented below each time samples change possession, including courier delivery.																
Relinquished by: 1 <u>Derek Jarman</u>	Date / Time: 4-9-21 1200	Received By: <u>[Signature]</u>	Date / Time: 4-12-21 0815	Relinquished By: 2	Date / Time: 4-12-21 0815	Received By: <u>[Signature]</u>	Date / Time: 4-12-21 0815	Relinquished By: 3	Date / Time: 4-12-21 0815	Received By: <u>[Signature]</u>	Date / Time: 4-12-21 0815	Relinquished By: 4	Date / Time: 4-12-21 0815	Received By: <u>[Signature]</u>	Date / Time: 4-12-21 0815	Relinquished By: 5
Relinquished by:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:
Relinquished by:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:
Relinquished by:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:
Relinquished by:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:
Relinquished by:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:	Date / Time:	Received By:	Date / Time:	Relinquished By:

31
3

TR 15

Signature

White Original: SGS copy Color Copy: Client copy

Copy of EHSA-QAC-0059-00-FORM-Houston - Air COC (5).xlsx
Rev. Date: 8/14/2018

JD23480: Chain of Custody

Page 1 of 2



SGS Sample Receipt Summary

Job Number: JD23480

Client: _____

Project: _____

Date / Time Received: 4/15/2021 10:30:00 AM

Delivery Method: _____

Airbill #'s: _____

Cooler Temps (Raw Measured) °C:

Cooler Temps (Corrected) °C:

Cooler Security

- | | |
|--|---|
| 1. Custody Seals Present: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | 3. COC Present: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> |
| 2. Custody Seals Intact: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | 4. Smpl Dates/Time OK: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> |

Cooler Temperature

- | | |
|--|-----|
| 1. Temp criteria achieved: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 2. Cooler temp verification: _____ | N/A |
| 3. Cooler media: _____ | N/A |
| 4. No. Coolers: _____ | N/A |

Quality Control Preservation

- | | |
|--|--|
| 1. Trip Blank present / cooler: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> <input checked="" type="checkbox"/> <u>N/A</u> | |
| 2. Trip Blank listed on COC: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> <input checked="" type="checkbox"/> <u>N/A</u> | |
| 3. Samples preserved properly: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 4. VOCs headspace free: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> <input checked="" type="checkbox"/> <u>N/A</u> | |

Sample Integrity - Documentation

- | | |
|---|--|
| 1. Sample labels present on bottles: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 2. Container labeling complete: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 3. Sample container label / COC agree: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |

Sample Integrity - Condition

- | | |
|---|--------|
| 1. Sample recvd within HT: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 2. All containers accounted for: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | |
| 3. Condition of sample: _____ | Intact |

Sample Integrity - Instructions

- | | | |
|---|--|--|
| 1. Analysis requested is clear: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | | |
| 2. Bottles received for unspecified tests: <input type="checkbox"/> <u>Y</u> <input checked="" type="checkbox"/> <u>N</u> | | |
| 3. Sufficient volume recvd for analysis: <input checked="" type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> | | |
| 4. Compositing instructions clear: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> <input checked="" type="checkbox"/> <u>N/A</u> | | |
| 5. Filtering instructions clear: <input type="checkbox"/> <u>Y</u> <input type="checkbox"/> <u>N</u> <input checked="" type="checkbox"/> <u>N/A</u> | | |

Test Strip Lot #s:	pH 1-12: _____	212820	pH 12+: _____	203117A	Other: (Specify) _____
--------------------	----------------	--------	---------------	---------	------------------------

Comments

SM089-03
Rev. Date 12/7/17

Summa Canister and Flow Controller Log

Job Number: JD23480
Account: PROVLABR Providence Engineering
Project: Valero-CAMS, Baton Rouge, LA
Received: 04/15/21

32
3

SUMMA CANISTERS													
Shipping							Receiving						
Summa ID	L	Vac " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
M009	6	29.4	03/18/21	WC	CP111072W52440.D		JD23480-1	04/16/21	JN				1

SGS Bottle Order(s):
 SW-031521-102

Prep Date **Room Temp(F)** **Bar Pres "Hg**
 03/18/21 70 29.92