

The results set forth herein are provided by SGS North America Inc.

e-Hardcopy 2.0
Automated Report

Technical Report for

Providence Engineering

Valero Refining Meraux, LA

712-001

SGS Job Number: JE31198

Sampling Date: 03/22/26

Report to:

Providence Engineering

1201 Main Street

Baton Rouge, LA 70802

brandonkilpatrick@providenceeng.com; daylansenecal@providenceeng.com

ATTN: Brandon Kilpatrick

Total number of pages in report: 35



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable unless noted in the narrative, comments or footnotes.

Olga Azarian
Technical Director

Client Service contact: Brittany Stacks 732-329-0200

Certifications: NJ(12129),NY(10983),CA,CO,CT,FL,HI,IL,IN,KY,LA (120428),MA,MD,ME,MN,NC,NH,NV,AK (UST-103),AZ (AZ0786),PA(68-00408),RI,SC,TX (T104704234),UT,VA,WA,WV

This report shall not be reproduced, except in its entirety, without the written approval of SGS.
Test results relate only to samples analyzed.

How did we do today?

Your feedback helps us improve our service and takes less than a minute to complete.

START SURVEY

Table of Contents

-1-

Section 1: Sample Summary	3
Section 2: Sample Results	4
2.1: JE31198-1: CAMS 917	5
Section 3: Misc. Forms	7
3.1: Chain of Custody	8
3.2: Summa Canister and Flow Controller Log	10
Section 4: MS Volatiles - QC Data Summaries	11
4.1: Method Blank Summary	12
4.2: Blank Spike/Blank Spike Duplicate Summary	18
4.3: Duplicate Summary	24
4.4: Summa Cleaning Certification	27
4.5: Instrument Performance Checks (BFB)	30
4.6: Surrogate Recovery Summaries	35



Sample Summary

Providence Engineering

Job No: JE31198

Valero Refining Meraux, LA
Project No: 712-001

Sample Number	Collected Date	Time By	Received	Matrix Code Type	Client Sample ID
---------------	----------------	---------	----------	------------------	------------------

This report contains results reported as ND = Not detected. The following applies:
Organics ND = Not detected above the MDL

JE31198-1	03/22/26	09:30 JD	03/31/26	AIR Ambient Air Comp.	CAMS 917
-----------	----------	----------	----------	-----------------------	----------

Sample Results

Report of Analysis

Report of Analysis

Client Sample ID: CAMS 917		
Lab Sample ID: JE31198-1		Date Sampled: 03/22/26
Matrix: AIR - Ambient Air Comp.	Summa ID: M044	Date Received: 03/31/26
Method: TO-15		Percent Solids: n/a
Project: Valero Refining Meraux, LA		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2W71195.D	1	04/01/26 21:19	TCH	n/a	n/a	V2W3229
Run #2							

Run #	Initial Volume
Run #1	400 ml
Run #2	

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
67-64-1	58.08	Acetone (2-Propanone)	4.6	0.20	0.15	ppbv		11	0.48	0.36	ug/m3
106-99-0	54.09	1,3-Butadiene	ND	0.20	0.084	ppbv		ND	0.44	0.19	ug/m3
71-43-2	78.11	Benzene	0.12	0.20	0.047	ppbv	J	0.38	0.64	0.15	ug/m3
75-27-4	163.8	Bromodichloromethane	ND	0.20	0.030	ppbv		ND	1.3	0.20	ug/m3
75-25-2	252.8	Bromoform	ND	0.20	0.071	ppbv		ND	2.1	0.73	ug/m3
74-83-9	94.94	Bromomethane	ND	0.20	0.069	ppbv		ND	0.78	0.27	ug/m3
593-60-2	106.9	Bromoethene	ND	0.20	0.061	ppbv		ND	0.87	0.27	ug/m3
100-44-7	126	Benzyl Chloride	ND	0.20	0.13	ppbv		ND	1.0	0.67	ug/m3
75-15-0	76.14	Carbon disulfide	ND	0.20	0.11	ppbv		ND	0.62	0.34	ug/m3
108-90-7	112.6	Chlorobenzene	ND	0.20	0.074	ppbv		ND	0.92	0.34	ug/m3
75-00-3	64.52	Chloroethane	ND	0.20	0.068	ppbv		ND	0.53	0.18	ug/m3
67-66-3	119.4	Chloroform	ND	0.20	0.037	ppbv		ND	0.98	0.18	ug/m3
74-87-3	50.49	Chloromethane	0.74	0.20	0.090	ppbv		1.5	0.41	0.19	ug/m3
107-05-1	76.53	3-Chloropropene	ND	0.20	0.083	ppbv		ND	0.63	0.26	ug/m3
95-49-8	126.6	2-Chlorotoluene	ND	0.20	0.072	ppbv		ND	1.0	0.37	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	0.040	ppbv		ND	1.3	0.25	ug/m3
110-82-7	84.16	Cyclohexane	ND	0.20	0.045	ppbv		ND	0.69	0.15	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	0.057	ppbv		ND	0.81	0.23	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	0.059	ppbv		ND	0.79	0.23	ug/m3
106-93-4	187.9	1,2-Dibromoethane (EDB)	ND	0.20	0.030	ppbv		ND	1.5	0.23	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	0.070	ppbv		ND	0.81	0.28	ug/m3
78-87-5	113	1,2-Dichloropropane	ND	0.20	0.062	ppbv		ND	0.92	0.29	ug/m3
123-91-1	88.12	1,4-Dioxane	ND	0.20	0.048	ppbv		ND	0.72	0.17	ug/m3
75-71-8	120.9	Dichlorodifluoromethane	0.43	0.20	0.10	ppbv		2.1	0.99	0.49	ug/m3
124-48-1	208.3	Dibromochloromethane	ND	0.20	0.12	ppbv		ND	1.7	1.0	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	0.069	ppbv		ND	0.79	0.27	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	0.030	ppbv		ND	0.79	0.12	ug/m3
10061-01-5	111	cis-1,3-Dichloropropene	ND	0.20	0.062	ppbv		ND	0.91	0.28	ug/m3
541-73-1	147	m-Dichlorobenzene	ND	0.20	0.028	ppbv		ND	1.2	0.17	ug/m3
95-50-1	147	o-Dichlorobenzene	ND	0.20	0.069	ppbv		ND	1.2	0.41	ug/m3
106-46-7	147	p-Dichlorobenzene	ND	0.20	0.079	ppbv		ND	1.2	0.47	ug/m3
10061-02-6	111	trans-1,3-Dichloropropene	ND	0.20	0.10	ppbv		ND	0.91	0.45	ug/m3

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	CAMS 917	Date Sampled:	03/22/26
Lab Sample ID:	JE31198-1	Date Received:	03/31/26
Matrix:	AIR - Ambient Air Comp. Summa ID: M044	Percent Solids:	n/a
Method:	TO-15		
Project:	Valero Refining Meraux, LA		

VOA TO15 List

CAS No.	MW	Compound	Result	RL	MDL	Units	Q	Result	RL	MDL	Units
64-17-5	46.07	Ethanol	11.7	0.50	0.39	ppbv		22.0	0.94	0.73	ug/m3
100-41-4	106.2	Ethylbenzene	ND	0.20	0.061	ppbv		ND	0.87	0.26	ug/m3
141-78-6	88	Ethyl Acetate	0.18	0.20	0.10	ppbv	J	0.65	0.72	0.36	ug/m3
622-96-8	120.19	4-Ethyltoluene	ND	0.20	0.095	ppbv		ND	0.98	0.47	ug/m3
76-13-1	187.4	Freon 113	ND	0.20	0.031	ppbv		ND	1.5	0.24	ug/m3
76-14-2	170.9	Freon 114	ND	0.20	0.050	ppbv		ND	1.4	0.35	ug/m3
142-82-5	100.2	Heptane	0.10	0.20	0.091	ppbv	J	0.41	0.82	0.37	ug/m3
87-68-3	260.8	Hexachlorobutadiene	ND	0.20	0.062	ppbv		ND	2.1	0.66	ug/m3
110-54-3	86.18	Hexane	0.67	0.20	0.052	ppbv		2.4	0.70	0.18	ug/m3
591-78-6	100	2-Hexanone	ND	0.20	0.15	ppbv		ND	0.82	0.61	ug/m3
67-63-0	60.1	Isopropyl Alcohol	0.46	0.20	0.14	ppbv		1.1	0.49	0.34	ug/m3
75-09-2	84.94	Methylene chloride	0.38	0.20	0.13	ppbv		1.3	0.69	0.45	ug/m3
78-93-3	72.11	Methyl ethyl ketone	2.3	0.20	0.052	ppbv		6.8	0.59	0.15	ug/m3
108-10-1	100.2	Methyl Isobutyl Ketone	ND	0.20	0.073	ppbv		ND	0.82	0.30	ug/m3
1634-04-4	88.15	Methyl Tert Butyl Ether	ND	0.20	0.080	ppbv		ND	0.72	0.29	ug/m3
80-62-6	100.12	Methylmethacrylate	ND	0.20	0.070	ppbv		ND	0.82	0.29	ug/m3
115-07-1	42	Propylene	ND	0.50	0.048	ppbv		ND	0.86	0.082	ug/m3
100-42-5	104.1	Styrene	ND	0.20	0.053	ppbv		ND	0.85	0.23	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	0.078	ppbv		ND	1.1	0.43	ug/m3
79-34-5	167.85	1,1,2,2-Tetrachloroethane	ND	0.20	0.048	ppbv		ND	1.4	0.33	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	0.038	ppbv		ND	1.1	0.21	ug/m3
120-82-1	181.5	1,2,4-Trichlorobenzene	ND	0.20	0.12	ppbv		ND	1.5	0.89	ug/m3
95-63-6	120.19	1,2,4-Trimethylbenzene	ND	0.20	0.087	ppbv		ND	0.98	0.43	ug/m3
108-67-8	120.19	1,3,5-Trimethylbenzene	ND	0.20	0.080	ppbv		ND	0.98	0.39	ug/m3
540-84-1	114.2	2,2,4-Trimethylpentane	ND	0.20	0.040	ppbv		ND	0.93	0.19	ug/m3
75-65-0	74.12	Tertiary Butyl Alcohol	ND	0.20	0.093	ppbv		ND	0.61	0.28	ug/m3
127-18-4	165.8	Tetrachloroethylene	0.038	0.040	0.030	ppbv	J	0.26	0.27	0.20	ug/m3
109-99-9	72.11	Tetrahydrofuran	17.7	0.20	0.090	ppbv		52.2	0.59	0.27	ug/m3
108-88-3	92.14	Toluene	0.27	0.20	0.057	ppbv		1.0	0.75	0.21	ug/m3
79-01-6	131.4	Trichloroethylene	ND	0.040	0.019	ppbv		ND	0.21	0.10	ug/m3
75-69-4	137.4	Trichlorofluoromethane	0.20	0.20	0.057	ppbv		1.1	1.1	0.32	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	0.069	ppbv		ND	0.51	0.18	ug/m3
108-05-4	86	Vinyl Acetate	0.39	0.20	0.11	ppbv		1.4	0.70	0.39	ug/m3
	106.2	m,p-Xylene	0.15	0.20	0.14	ppbv	J	0.65	0.87	0.61	ug/m3
95-47-6	106.2	o-Xylene	ND	0.20	0.077	ppbv		ND	0.87	0.33	ug/m3
1330-20-7	106.2	Xylenes (total)	0.15	0.20	0.077	ppbv	J	0.65	0.87	0.33	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	102%		65-128%

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody
- Summa Canister and Flow Controller Log



Air

AIR CHAIN OF CUSTODY

PAGE 0F

SGS North America Inc. - Dayton
2235 Route 130, Dayton, NJ 08810
TEL 732-329-0200 FAX 732-329-3499
www.sgs.com/ehsusa

Form Title: **8091897064701** Boiler Order Control # **55-122721-154**
SGS Quote # _____
SGS Job # **JE31198**

Client / Reporting Information				Project Information				Weather Parameters				Requested Analysis							
Company Name: Providence Engineering				Project Name: Valero refining				Temperature (Fahrenheit)				Requested Analysis							
Address: 1201 Main Street				Street				Start: Maximum:											
City: Burton Rouse State: CA Zip: 70802				City: Meroux State: LA				Stop: Minimum:											
Project Contact: Brandon Kipatruck E-mail: brandon.kipatruck@providence.org				Project # 712-001				Atmospheric Pressure (inches of Hg)											
Phone # 215-766-7400 Fax # _____				Client Purchase Order # _____				Start: Maximum:											
Sampler(s) Name(s)				Other weather comment:				Stop: Minimum:											
Lab Sample #	Field ID / Point of Collection	Air Type	Sampling Equipment Info			Start Sampling Information					Stop Sampling Information								
		Indoor (I) / Soil Vap (SV) / Ambient (A)	Canister Serial #	Canister Size 6L or 1L	Flow Controller Serial #	Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.	Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.				
1	Cans 917	A	10844	6L	687	3/22/24	9:30 am	30	60	6C	3/22	9:30 am	4	70	JD	X			
Turnaround Time (Business days)				Data Deliverable Information				Comments / Remarks											
<input checked="" type="checkbox"/> Standard - 15 Days <input type="checkbox"/> 10 Day <input type="checkbox"/> 5 Day <input type="checkbox"/> 3 Day <input type="checkbox"/> 2 Day <input type="checkbox"/> 1 Day <input type="checkbox"/> Other				Approved By: _____ Date: _____				All NJDEP TO-15 is mandatory Full T1 Comm A _____ Comm B _____ Reduced T2 _____ Full T1 _____ Other: _____ DKQP reporting _____				Sample Assessment AR4B Sample inventory is verified upon receipt in the Laboratory							
Sample Custody must be documented below each time samples change possession, including courier delivery.																			
Relinquished by: _____		Date/Time: 3/21/24		Received By: _____		Date/Time: _____		Relinquished By: _____		Date/Time: _____		Received By: _____		Date/Time: _____		Received By: _____			
Relinquished by: _____		Date/Time: 3/23/24		Received By: _____		Date/Time: _____		Relinquished By: _____		Date/Time: _____		Received By: _____		Date/Time: _____		Received By: _____			
Relinquished by: _____		Date/Time: _____		Received By: _____		Date/Time: _____		Relinquished By: _____		Date/Time: _____		Received By: _____		Date/Time: _____		Received By: _____			
Relinquished by: _____		Date/Time: _____		Received By: _____		Date/Time: _____		Relinquished By: _____		Date/Time: _____		Received By: _____		Date/Time: _____		Received By: _____			

31
3

